

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD011222AL\
 Data File : FD040563.D
 Signal(s) : FID2B.CH
 Acq On : 12 Jan 2022 9:50
 Operator : AJ\MA
 Sample : 20 PPM ALIPHATIC HC STD
 Misc :
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 FID_D
 LabSampleId :
 20 PPM ALIPHATIC HC STD

Integration File: autoint1.e
 Quant Time: Jan 13 01:17:08 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aliphatic EPH 010422.M
 Quant Title : GC Extractables
 QLast Update : Wed Jan 05 07:08:39 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 T n-Nonane (C9)	128.382	133.019 E3	-3.6	102	0.00
2 T n-Decane (C10)	128.766	135.579 E3	-5.3	103	0.00
3 T A~Naphthalene (C11.7)	142.072	150.835 E3	-6.2	103	0.00
4 T n-Dodecane (C12)	127.814	137.755 E3	-7.8	106	0.00
5 T A~2-methylnaphthalene (C12.	141.074	152.543 E3	-8.1	105	0.00
6 T n-Tetradecane (C14)	129.540	141.602 E3	-9.3	107	0.00
7 T n-Hexadecane (C16)	130.284	142.584 E3	-9.4	107	0.00
8 T n-Octadecane (C18)	135.501	145.253 E3	-7.2	105	0.00
9 S ortho-Terphenyl (SURR)	154.188	158.162 E3	-2.6	101	-0.01
10 T n-Eicosane (C20)	136.020	143.357 E3	-5.4	104	-0.01
11 T n-Heneicosane (C21)	138.476	141.289 E3	-2.0	100	-0.01
12 S 1-chlorooctadecane (SURR)	117.707	119.689 E3	-1.7	99	-0.01
13 T n-Docosane (C22)	137.887	145.706 E3	-5.7	103	-0.01
14 T n-Tetracosane (C24)	139.825	153.993 E3	-10.1	108	-0.01
15 T n-Hexacosane (C26)	137.845	154.109 E3	-11.8	109	-0.01
16 T n-Octacosane (C28)	135.686	155.665 E3	-14.7	112	-0.01
17 T n-Tricontane (C30)	130.338	153.273 E3	-17.6	115	-0.01
18 T n-Dotriacontane (C32)	119.545	140.248 E3	-17.3	114	-0.01
19 T n-Tetratriacontane (C34)	105.292	122.935 E3	-16.8	116	-0.01
20 T n-Hexatriacontane (C36)	101.331	106.343 E3	-4.9	106	-0.01
21 T n-Octatriacontane (C38)	89.345	97.577 E3	-9.2	107	-0.01
22 T n-Tetracontane (C40)	85.482	96.562 E3	-13.0	108	-0.01

Evaluate Continuing Calibration Report - Not Found

(#) = Out of Range

SPCC's out = 0 CCC's out = 0