

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD011421AL\
 Data File : FD036040.D
 Signal(s) : FID2B.CH
 Acq On : 14 Jan 2021 16:42
 Operator : DD\AJ
 Sample : M1121-02
 Misc :
 ALS Vial : 57 Sample Multiplier: 1

Instrument :
 FID_D
 ClientSampleId :

Integration File: autoint1.e
 Quant Time: Jan 14 21:26:08 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Alaphatic EPH 011121.M
 Quant Title : GC Extractables
 QLast Update : Tue Jan 12 02:11:59 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.531	2088147	14.087 ug/ml
Spiked Amount 50.000		Recovery =	28.17%
12) S 1-chlorooctadecane (S...	12.965	1503436	12.781 ug/ml
Spiked Amount 50.000		Recovery =	25.56%
Target Compounds			
6) T n-Tetradecane (C14)	8.149	203453	1.523 ug/ml
10) T n-Eicosane (C20)	12.571f	166509	1.205 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD011421AL\
Data File : FD036040.D
Signal(s) : FID2B.CH
Acq On : 14 Jan 2021 16:42
Operator : DD\AJ
Sample : M1121-02
Misc :
ALS Vial : 57 Sample Multiplier: 1

Instrument :
FID_D
ClientSampleId :

Integration File: autoint1.e
Quant Time: Jan 14 21:26:08 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Alaphatic EPH 011121.M
Quant Title : GC Extractables
QLast Update : Tue Jan 12 02:11:59 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#6 n-Tetradecane (C14)


R.T.: 8.149 min
Delta R.T.: -0.030 min
Response: 203453
Conc: 1.52 ug/ml

Instrument :
FID_D
ClientSampleId :




#9 ortho-Terphenyl (SURR)

R.T.: 11.531 min
Delta R.T.: 0.009 min
Response: 2088147
Conc: 14.09 ug/ml



#10 n-Eicosane (C20)

R.T.: 12.571 min
Delta R.T.: 0.057 min
Response: 166509
Conc: 1.21 ug/ml



#12 1-chlorooctadecane (SURR)

R.T.: 12.965 min
Delta R.T.: 0.009 min
Response: 1503436
Conc: 12.78 ug/ml