

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD012320AL\
 Data File : FD032126.D
 Signal(s) : FID2B.CH
 Acq On : 23 Jan 2020 22:46
 Operator : DD\AJ
 Sample : L1239-08DL 2X
 Misc :
 ALS Vial : 69 Sample Multiplier: 1

Instrument :
 FID_D
 ClientSampleId :

Integration File: autoint1.e
 Quant Time: Jan 24 02:53:55 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Alaphatic EPH 010620.M
 Quant Title : GC Extractables
 QLast Update : Tue Jan 07 01:05:13 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.535	2274215	16.462 ug/ml
Spiked Amount 50.000		Recovery =	32.92%
12) S 1-chlorooctadecane (S...	12.966	1764477	16.198 ug/ml
Spiked Amount 50.000		Recovery =	32.40%
Target Compounds			
6) T n-Tetradecane (C14)	8.249f	198374	1.634 ug/ml
10) T n-Eicosane (C20)	12.579	414839	3.251 ug/ml
11) T n-Heneicosane (C21)	13.191f	384539	2.985 ug/ml
16) T n-Octacosane (C28)	16.824	2828767	22.192 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD012320AL\
Data File : FD032126.D
Signal(s) : FID2B.CH
Acq On : 23 Jan 2020 22:46
Operator : DD\AJ
Sample : L1239-08DL 2X
Misc :
ALS Vial : 69 Sample Multiplier: 1

Instrument :
FID_D
ClientSampleId :

Integration File: autoint1.e
Quant Time: Jan 24 02:53:55 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Alaphatic EPH 010620.M
Quant Title : GC Extractables
QLast Update : Tue Jan 07 01:05:13 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#6 n-Tetradecane (C14)


R.T.: 8.249 min
Delta R.T.: 0.050 min
Response: 198374
Conc: 1.63 ug/ml

Instrument :
FID_D
ClientSampleId :




#9 ortho-Terphenyl (SURR)

R.T.: 11.535 min
Delta R.T.: -0.009 min
Response: 2274215
Conc: 16.46 ug/ml




#10 n-Eicosane (C20)

R.T.: 12.579 min
Delta R.T.: 0.049 min
Response: 414839
Conc: 3.25 ug/ml



#11 n-Heneicosane (C21)


R.T.: 13.191 min
Delta R.T.: 0.053 min
Response: 384539
Conc: 2.99 ug/ml



#12 1-chlorooctadecane (SURR)

R.T.: 12.966 min
Delta R.T.: -0.006 min
Response: 1764477
Conc: 16.20 ug/ml

Instrument :
FID_D
ClientSampleId :



#16 n-Octacosane (C28)

R.T.: 16.824 min
Delta R.T.: 0.047 min
Response: 2828767
Conc: 22.19 ug/ml