

Demonstration of Capability (DoC)

System : GENCHEM Matrix : WATER
Analysis Method : SM3500-Cr-B Analyst : Indiana Benegas

Analyte	Replicate Concentrations	Units: mg/L		Average Recovery (%R)	Acceptance Window (%R)	RSD Acceptance	
		LB99503BS	LB99466BS LB99669BS			RSD	Limit
Hexavalent Chromium	98.8	97.8	97	99	98.15	90-111	1 20