

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD020719AR\
 Data File : FD027639.D
 Signal(s) : FID2B.CH
 Acq On : 08 Feb 2019 3:55
 Operator : AJ\SJ
 Sample : 20 PPM AROMATIC HC STD
 Misc :
 ALS Vial : 52 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Feb 08 08:03:12 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aromatic EPH 020419.M
 Quant Title : GC Extractables
 QLast Update : Tue Feb 05 01:03:41 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 T	1,2,3-Trimethylbenzene (C10	79.035	83.471 E3	-5.6	104	0.00
2 T	Naphthalene (C11.7)	85.118	88.930 E3	-4.5	103	0.00
3 T	2-Methylnaphthalene (C12.89	84.365	88.651 E3	-5.1	103	0.00
4 S	2-Bromonaphthalene (SURR)	74.240	78.013 E3	-5.1	102	0.00
5 T	Acenaphthylene (C15.06)	83.729	87.542 E3	-4.6	102	0.00
6 S	2-Fluorobiphenyl (SURR)	50.878	53.415 E3	-5.0	101	0.00
7 T	Acenaphthene (C15.5)	87.899	91.807 E3	-4.4	102	0.00
8 T	Flouorene (C16.55)	87.209	90.478 E3	-3.7	101	0.00
9 T	Phenanthrene (C19.36)	87.843	90.804 E3	-3.4	100	0.00
10 T	Anthracene (C19.43)	86.409	89.688 E3	-3.8	101	0.00
11 S	ortho-Terphenyl (SURR)	88.645	93.622 E3	-5.6	103	0.00
12 T	Fluoranthene (C21.85)	89.526	93.289 E3	-4.2	103	0.00
13 T	Pyrene (C20.8)	88.736	93.551 E3	-5.4	103	0.00
14 T	Benzo[a]anthracene (C26.37)	89.416	93.051 E3	-4.1	103	0.00
15 T	Chrysene (C27.41)	90.813	94.410 E3	-4.0	104	0.00
16 T	benzo[b]fluoranthene (C30.4	97.666	96.526 E3	1.2	100	0.00
17 T	Bnezo[k]fluoranthene (C30.1	98.876	98.150 E3	0.7	101	0.00
18 T	Benzo[a]pyrene (C31.34)	97.777	95.852 E3	2.0	98	0.00
19 T	Indeno[1,2,3-cd]pyrene (C35	99.167	93.944 E3	5.3	95	0.00
20 T	Dibenz[a,h]anthracene (C30.	98.362	93.248 E3	5.2	93	0.00
21 T	Benzo[g,h,i]perylene (C34.0	101.380	95.692 E3	5.6	96	0.00

Evaluate Continuing Calibration Report - Not Founds

(#) = Out of Range

SPCC's out = 0 CCC's out = 0