

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD020819AR\
 Data File : FD027648.D
 Signal(s) : FID2B.CH
 Acq On : 08 Feb 2019 13:54
 Operator : AJ\SJ
 Sample : 20 PPM AROMATIC HC STD
 Misc :
 ALS Vial : 52 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Feb 08 16:19:20 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aromatic EPH 020419.M
 Quant Title : GC Extractables
 QLast Update : Tue Feb 05 01:03:41 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 T	1,2,3-Trimethylbenzene (C10	79.035	88.159 E3	-11.5	110	0.00
2 T	Naphthalene (C11.7)	85.118	94.468 E3	-11.0	109	0.00
3 T	2-Methylnaphthalene (C12.89	84.365	92.963 E3	-10.2	108	0.00
4 S	2-Bromonaphthalene (SURR)	74.240	82.075 E3	-10.6	108	0.00
5 T	Acenaphthylene (C15.06)	83.729	89.198 E3	-6.5	104	0.00
6 S	2-Fluorobiphenyl (SURR)	50.878	54.217 E3	-6.6	103	0.00
7 T	Acenaphthene (C15.5)	87.899	93.477 E3	-6.3	104	0.00
8 T	Flouorene (C16.55)	87.209	90.676 E3	-4.0	101	0.00
9 T	Phenanthrene (C19.36)	87.843	90.277 E3	-2.8	100	0.00
10 T	Anthracene (C19.43)	86.409	89.025 E3	-3.0	100	0.00
11 S	ortho-Terphenyl (SURR)	88.645	93.739 E3	-5.7	103	0.00
12 T	Fluoranthene (C21.85)	89.526	94.378 E3	-5.4	104	0.00
13 T	Pyrene (C20.8)	88.736	94.691 E3	-6.7	104	0.00
14 T	Benzo[a]anthracene (C26.37)	89.416	95.069 E3	-6.3	105	0.00
15 T	Chrysene (C27.41)	90.813	96.403 E3	-6.2	106	0.00
16 T	benzo[b]fluoranthene (C30.4	97.666	98.414 E3	-0.8	102	0.00
17 T	Bnezo[k]fluoranthene (C30.1	98.876	101.050 E3	-2.2	104	0.00
18 T	Benzo[a]pyrene (C31.34)	97.777	97.603 E3	0.2	100	0.00
19 T	Indeno[1,2,3-cd]pyrene (C35	99.167	95.975 E3	3.2	97	0.00
20 T	Dibenz[a,h]anthracene (C30.	98.362	95.144 E3	3.3	95	0.00
21 T	Benzo[g,h,i]perylene (C34.0	101.380	97.747 E3	3.6	98	0.00

Evaluate Continuing Calibration Report - Not Founds

(#) = Out of Range

SPCC's out = 0 CCC's out = 0