

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD021323AR\
 Data File : FD045016.D
 Signal(s) : FID2B.ch
 Acq On : 13 Feb 2023 15:34
 Operator : YP/AJ
 Sample : 20 PPM AROMATIC HC STD
 Misc :
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 FID_D
 ClientSampleId :
 20 PPM AROMATIC HC STD

Manual Integrations
 APPROVED

Reviewed By :Ankita Jodhani 02/14/2023
 Supervised By :Sohil Jodhani 02/14/2023

Integration File: autoint1.e
 Quant Time: Feb 14 00:27:11 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aromatic EPH 012323.M
 Quant Title : GC Extractables
 QLast Update : Mon Jan 23 15:00:18 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18µm

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
4) S 2-Bromonaphthalene (S...	7.556	2832329	19.312 ug/ml
Spiked Amount 50.000		Recovery =	38.62%
6) S 2-Fluorobiphenyl (SURR)	8.413	1898461	19.020 ug/ml
Spiked Amount 50.000 Range 0 - 131		Recovery =	38.04%
11) S ortho-Terphenyl (SURR)	11.455	3588827	19.309 ug/ml
Spiked Amount 50.000		Recovery =	38.62%
Target Compounds			
1) T 1,2,3-Trimethylbenzen...	4.361	3143597	19.693 ug/ml
2) T Naphthalene (C11.7)	5.894	3323031	19.613 ug/ml
3) T 2-Methylnaphthalene (...)	6.944	3294044	19.201 ug/ml
5) T Acenaphthylene (C15.06)	8.217	3021869	19.230 ug/ml
7) T Acenaphthene (C15.5)	8.513	3201215	19.253 ug/ml
8) T Flouorene (C16.55)	9.292	3117169	18.991 ug/ml
9) T Phenanthrene (C19.36)	10.684	3079210	19.045 ug/ml
10) T Anthracene (C19.43)	10.758	2830998	19.096 ug/ml
12) T Fluoranthene (C21.85)	12.492	3075696	19.218 ug/ml
13) T Pyrene (C20.8)	12.787	2995955	19.093 ug/ml
14) T Benzo[a]anthracene (C...	14.652	2695205	18.884 ug/ml
15) T Chrysene (C27.41)	14.695	2812571	18.996 ug/ml
16) T benzo[b]fluoranthene ...	16.197	2810326	19.372 ug/ml
17) T Bnezo[k]fluoranthene ...	16.230	2823781	19.806 ug/ml
18) T Benzo[a]pyrene (C31.34)	16.574	2747824	19.664 ug/ml
19) T Indeno[1,2,3-cd]pyren...	17.932	2767429	20.565 ug/mlm
20) T Dibenz[a,h]anthracene...	17.974	2731229	20.610 ug/ml
21) T Benzo[g,h,i]perylene ...	18.187	2745802	20.195 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD021323AR\
Data File : FD045016.D
Signal(s) : FID2B.ch
Acq On : 13 Feb 2023 15:34
Operator : YP/AJ
Sample : 20 PPM AROMATIC HC STD
Misc :
ALS Vial : 52 Sample Multiplier: 1

Instrument :
FID_D
ClientSampleId :
20 PPM AROMATIC HC STD

Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 02/14/2023
Supervised By :Sohil Jodhani 02/14/2023

Integration File: autoint1.e
Quant Time: Feb 14 00:27:11 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aromatic EPH 012323.M
Quant Title : GC Extractables
QLast Update : Mon Jan 23 15:00:18 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18µm

