

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_D\Data\FD022122AL\  
 Data File : FD041219.D  
 Signal(s) : FID2B.CH  
 Acq On : 21 Feb 2022 17:40  
 Operator : YP\AJ  
 Sample : N1589-02D  
 Misc :  
 ALS Vial : 60 Sample Multiplier: 1

Instrument :  
 FID\_D  
 ClientSampleId :

Integration File: autoint1.e  
 Quant Time: Feb 22 02:18:08 2022  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_D\methods\Aliphatic EPH 020822.M  
 Quant Title : GC Extractables  
 QLast Update : Wed Feb 09 03:32:15 2022  
 Response via : Initial Calibration  
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul  
 Signal Phase : Rxi-1ms  
 Signal Info : 20M x 0.18mm x 0.18um

| Compound                       | R.T.   | Response   | Conc Units   |
|--------------------------------|--------|------------|--------------|
| -----                          |        |            |              |
| System Monitoring Compounds    |        |            |              |
| 9) S ortho-Terphenyl (SURR)    | 11.230 | 1713055    | 61.291 ug/ml |
| Spiked Amount 50.000           |        | Recovery = | 122.58%      |
| 12) S 1-chlorooctadecane (S... | 12.671 | 1232005    | 58.849 ug/ml |
| Spiked Amount 50.000           |        | Recovery = | 117.70%      |
| Target Compounds               |        |            |              |
| 10) T n-Eicosane (C20)         | 12.260 | 55303      | 2.208 ug/ml  |
| -----                          |        |            |              |

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_D\Data\FD022122AL\  
Data File : FD041219.D  
Signal(s) : FID2B.CH  
Acq On : 21 Feb 2022 17:40  
Operator : YP\AJ  
Sample : N1589-02D  
Misc :  
ALS Vial : 60 Sample Multiplier: 1

Instrument :  
FID\_D  
ClientSampleId :

Integration File: autoint1.e  
Quant Time: Feb 22 02:18:08 2022  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_D\methods\Aliphatic EPH 020822.M  
Quant Title : GC Extractables  
QLast Update : Wed Feb 09 03:32:15 2022  
Response via : Initial Calibration  
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul  
Signal Phase : Rxi-1ms  
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)


R.T.: 11.230 min  
Delta R.T.: -0.005 min  
Response: 1713055  
Conc: 61.29 ug/ml

Instrument :  
FID\_D  
ClientSampleId :



#10 n-Eicosane (C20)

R.T.: 12.260 min  
Delta R.T.: 0.021 min  
Response: 55303  
Conc: 2.21 ug/ml



#12 1-chlorooctadecane (SURR)

R.T.: 12.671 min  
Delta R.T.: -0.006 min  
Response: 1232005  
Conc: 58.85 ug/ml