

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD022221AL\
 Data File : FD036245.D
 Signal(s) : FID2B.CH
 Acq On : 22 Feb 2021 20:08
 Operator : DD\AJ
 Sample : M1452-09
 Misc :
 ALS Vial : 66 Sample Multiplier: 1

Instrument :
 FID_D
 ClientSampleId :

Integration File: autoint1.e
 Quant Time: Feb 23 03:57:11 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Alaphatic EPH 021521.M
 Quant Title : GC Extractables
 QLast Update : Tue Feb 16 04:34:45 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.512	5606764	40.608 ug/ml
Spiked Amount 50.000		Recovery =	81.22%
12) S 1-chlorooctadecane (S...	12.944	4381107	39.778 ug/ml
Spiked Amount 50.000		Recovery =	79.56%
Target Compounds			
10) T n-Eicosane (C20)	12.547	664667	5.146 ug/ml
14) T n-Tetracosane (C24)	14.751	399521	3.055 ug/ml
16) T n-Octacosane (C28)	16.736	110500	0.848 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD022221AL\
Data File : FD036245.D
Signal(s) : FID2B.CH
Acq On : 22 Feb 2021 20:08
Operator : DD\AJ
Sample : M1452-09
Misc :
ALS Vial : 66 Sample Multiplier: 1

Instrument :
FID_D
ClientSampleId :

Integration File: autoint1.e
Quant Time: Feb 23 03:57:11 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Alaphatic EPH 021521.M
Quant Title : GC Extractables
QLast Update : Tue Feb 16 04:34:45 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)


R.T.: 11.512 min
Delta R.T.: 0.002 min
Response: 5606764
Conc: 40.61 ug/ml

Instrument :
FID_D
ClientSampleId :




#10 n-Eicosane (C20)

R.T.: 12.547 min
Delta R.T.: 0.047 min
Response: 664667
Conc: 5.15 ug/ml




#12 1-chlorooctadecane (SURR)

R.T.: 12.944 min
Delta R.T.: 0.003 min
Response: 4381107
Conc: 39.78 ug/ml



#14 n-Tetracosane (C24)

R.T.: 14.751 min
Delta R.T.: -0.038 min
Response: 399521
Conc: 3.05 ug/ml



#16 n-Octacosane (C28)

R.T.: 16.736 min
Delta R.T.: -0.012 min
Response: 110500
Conc: 0.85 ug/ml

Instrument :
FID_D
ClientSampleId :