

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD022520AR\
 Data File : FD032420.D
 Signal(s) : FID2B.CH
 Acq On : 25 Feb 2020 9:36
 Operator : DD\AJ
 Sample : 20 PPM AROMATIC HC STD
 Misc :
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 FID_D
 LabSampleId :
 20 PPM AROMATIC HC STD

Integration File: autoint1.e
 Quant Time: Feb 26 00:30:07 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aromatic EPH 020420.M
 Quant Title : GC Extractables
 QLast Update : Wed Feb 05 11:53:53 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 T	1,2,3-Trimethylbenzene (C10	144.821	142.450 E3	1.6	102	0.00
2 T	Naphthalene (C11.7)	152.396	147.108 E3	3.5	100	0.00
3 T	2-Methylnaphthalene (C12.89	151.284	143.563 E3	5.1	98	0.00
4 S	2-Bromonaphthalene (SURR)	138.969	130.240 E3	6.3	97	0.00
5 T	Acenaphthylene (C15.06)	147.353	132.327 E3	10.2	92	0.00
6 S	2-Fluorobiphenyl (SURR)	94.833	87.377 E3	7.9	94	0.00
7 T	Acenaphthene (C15.5)	151.013	139.114 E3	7.9	94	0.00
8 T	Flouorene (C16.55)	149.658	135.589 E3	9.4	92	0.00
9 T	Phenanthrene (C19.36)	151.242	131.327 E3	13.2	88	-0.01
10 T	Anthracene (C19.43)	148.332	117.906 E3	20.5	81	0.00
11 S	ortho-Terphenyl (SURR)	148.327	130.637 E3	11.9	90	0.00
12 T	Fluoranthene (C21.85)	150.585	125.860 E3	16.4	85	-0.01
13 T	Pyrene (C20.8)	149.287	123.918 E3	17.0	84	-0.01
14 T	Benzo[a]anthracene (C26.37)	142.018	115.066 E3	19.0	82	-0.01
15 T	Chrysene (C27.41)	142.950	118.185 E3	17.3	84	-0.01
16 T	benzo[b]fluoranthene (C30.4	138.579	116.114 E3	16.2	85	-0.02
17 T	Bnezo[k]fluoranthene (C30.1	138.016	115.908 E3	16.0	85	-0.02
18 T	Benzo[a]pyrene (C31.34)	136.246	114.638 E3	15.9	85	-0.02
19 T	Indeno[1,2,3-cd]pyrene (C35	129.348	112.941 E3	12.7	88	-0.02
20 T	Dibenz[a,h]anthracene (C30.	130.536	117.065 E3	10.3	90	-0.02
21 T	Benzo[g,h,i]perylene (C34.0	125.593	113.787 E3	9.4	91	-0.02

Evaluate Continuing Calibration Report - Not Founds

(#) = Out of Range

SPCC's out = 0 CCC's out = 0