

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD022719AL\
 Data File : FD027990.D
 Signal(s) : FID2B.CH
 Acq On : 28 Feb 2019 00:40
 Operator : AJ\SJ
 Sample : 20 PPM ALIPHATIC HC STD
 Misc :
 ALS Vial : 62 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Feb 28 08:57:22 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Alaphatic EPH 020419.M
 Quant Title : GC Extractables
 QLast Update : Tue Feb 05 08:23:12 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 T n-Nonane (C9)	82.001	91.869 E3	-12.0	115	0.00
2 T n-Decane (C10)	82.325	93.634 E3	-13.7	117	0.00
3 T A~Naphthalene (C11.7)	93.146	103.360 E3	-11.0	114	-0.01
4 T n-Dodecane (C12)	82.173	94.320 E3	-14.8	119	0.00
5 T A~2-methylnaphthalene (C12.	93.332	103.182 E3	-10.6	113	-0.01
6 T n-Tetradecane (C14)	84.037	93.651 E3	-11.4	116	-0.01
7 T n-Hexadecane (C16)	84.860	91.539 E3	-7.9	113	-0.01
8 T n-Octadecane (C18)	87.113	90.068 E3	-3.4	109	-0.01
9 S ortho-Terphenyl (SURR)	114.882	117.236 E3	-2.0	105	-0.01
10 T n-Eicosane (C20)	85.873	89.171 E3	-3.8	107	-0.02
11 T n-Heneicosane (C21)	84.267	88.332 E3	-4.8	107	-0.02
12 S 1-chlorooctadecane (SURR)	85.369	84.126 E3	1.5	101	-0.02
13 T n-Docosane (C22)	84.450	87.821 E3	-4.0	106	-0.02
14 T n-Tetracosane (C24)	84.726	87.762 E3	-3.6	105	-0.02
15 T n-Hexacosane (C26)	84.477	87.261 E3	-3.3	104	-0.02
16 T n-Octacosane (C28)	85.127	87.631 E3	-2.9	104	-0.02
17 T n-Tricontane (C30)	84.239	86.538 E3	-2.7	104	-0.02
18 T n-Dotriacontane (C32)	82.503	83.382 E3	-1.1	103	-0.02
19 T n-Tetratriacontane (C34)	72.494	72.545 E3	-0.1	102	-0.02
20 T n-Hexatriacontane (C36)	63.365	61.518 E3	2.9	98	-0.02
21 T n-Octatriacontane (C38)	59.431	53.434 E3	10.1	91	-0.02
22 T n-Tetracontane (C40)	57.281	44.214 E3	22.8	77	-0.03

Evaluate Continuing Calibration Report - Not Found

(#) = Out of Range

SPCC's out = 0 CCC's out = 0