

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_D\Data\FD022719AR\  
Data File : FD027987.D  
Signal(s) : FID2B.CH  
Acq On : 27 Feb 2019 22:52  
Operator : AJ\SJ  
Sample : I.BLK  
Misc :  
ALS Vial : 51 Sample Multiplier: 1

Integration File: autoint1.e  
Quant Time: Feb 28 03:07:43 2019  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_D\methods\Aromatic EPH 020419.M  
Quant Title : GC Extractables  
QLast Update : Tue Feb 05 01:03:41 2019  
Response via : Initial Calibration  
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl  
Signal Phase : Rxi-1ms  
Signal Info : 20M x 0.18mm x 0.18µm

| Compound | R.T. | Response | Conc Units |
|----------|------|----------|------------|
|----------|------|----------|------------|

-----  
System Monitoring Compounds

Target Compounds

|       |                          |         |        |             |
|-------|--------------------------|---------|--------|-------------|
| 14) T | Benzo[a]anthracene (C... | 14.346f | 220424 | 2.465 ug/ml |
| 16) T | benzo[b]fluoranthene ... | 16.000  | 109145 | 1.118 ug/ml |
| 17) T | Bnezo[k]fluoranthene ... | 16.000  | 109145 | 1.104 ug/ml |
| 19) T | Indeno[1,2,3-cd]pyren... | 17.763f | 158327 | 1.597 ug/ml |
| 20) T | Dibenz[a,h]anthracene... | 17.763  | 158327 | 1.610 ug/ml |


-----  
(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_D\Data\FD022719AR\  
Data File : FD027987.D  
Signal(s) : FID2B.CH  
Acq On : 27 Feb 2019 22:52  
Operator : AJ\SJ  
Sample : I.BLK  
Misc :  
ALS Vial : 51 Sample Multiplier: 1

Integration File: autoint1.e  
Quant Time: Feb 28 03:07:43 2019  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_D\methods\Aromatic EPH 020419.M  
Quant Title : GC Extractables  
QLast Update : Tue Feb 05 01:03:41 2019  
Response via : Initial Calibration  
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped


Volume Inj. : 1 µl  
Signal Phase : Rxi-1ms  
Signal Info : 20M x 0.18mm x 0.18µm






#4 2-Bromonaphthalene (SURRE)

R.T.: 0.000 min  
Exp R.T. : 7.345 min  
Response: 0  
Conc: N.D.




#6 2-Fluorobiphenyl (SURRE)

R.T.: 0.000 min  
Exp R.T. : 8.194 min  
Response: 0  
Conc: N.D.




#11 ortho-Terphenyl (SURRE)

R.T.: 0.000 min  
Exp R.T. : 11.231 min  
Response: 0  
Conc: N.D.




#14 Benzo[a]anthracene (C26.37)

R.T.: 14.346 min  
Delta R.T.: -0.070 min  
Response: 220424  
Conc: 2.47 ug/ml




#16 benzo[b]fluoranthene (C30.41)

R.T.: 16.000 min  
Delta R.T.: 0.042 min  
Response: 109145  
Conc: 1.12 ug/ml




#17 Bnezo[k]fluoranthene (C30.14)

R.T.: 16.000 min  
Delta R.T.: 0.010 min  
Response: 109145  
Conc: 1.10 ug/ml



#19 Indeno[1,2,3-cd]pyrene (C35.01)

R.T.: 17.763 min  
Delta R.T.: 0.070 min  
Response: 158327  
Conc: 1.60 ug/ml



#20 Dibenz[a,h]anthracene (C30.36)

R.T.: 17.763 min  
Delta R.T.: 0.029 min  
Response: 158327  
Conc: 1.61 ug/ml