

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD022823AR\
 Data File : FD045189.D
 Signal(s) : FID2B.ch
 Acq On : 28 Feb 2023 22:07
 Operator : YP/AJ
 Sample : 20 PPM AROMATIC HC STD
 Misc :
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 FID_D
 ClientSampleId :
 20 PPM AROMATIC HC STD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 03/01/2023
 Supervised By :Ankita Jodhani 03/01/2023

Integration File: autoint1.e
 Quant Time: Mar 01 04:21:16 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aromatic EPH 022123.M
 Quant Title : GC Extractables
 QLast Update : Wed Feb 22 04:55:52 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18µm

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
4) S 2-Bromonaphthalene (S...	7.552	2790672	20.807 ug/ml
Spiked Amount 50.000		Recovery =	41.61%
6) S 2-Fluorobiphenyl (SURR)	8.408	1855064	20.699 ug/ml
Spiked Amount 50.000 Range 0 - 131		Recovery =	41.40%
11) S ortho-Terphenyl (SURR)	11.449	3475681	20.665 ug/ml
Spiked Amount 50.000		Recovery =	41.33%
Target Compounds			
1) T 1,2,3-Trimethylbenzen...	4.361	3126662	20.885 ug/ml
2) T Naphthalene (C11.7)	5.892	3269063	20.821 ug/ml
3) T 2-Methylnaphthalene (...)	6.941	3254416	20.802 ug/ml
5) T Acenaphthylene (C15.06)	8.212	2958686	20.722 ug/ml
7) T Acenaphthene (C15.5)	8.507	3143350	20.845 ug/ml
8) T Fluorene (C16.55)	9.286	3046760	20.670 ug/ml
9) T Phenanthrene (C19.36)	10.678	2957151	20.428 ug/ml
10) T Anthracene (C19.43)	10.751	2577948	19.041 ug/ml
12) T Fluoranthene (C21.85)	12.482	2916846	20.350 ug/ml
13) T Pyrene (C20.8)	12.777	2843564	20.307 ug/ml
14) T Benzo[a]anthracene (C...	14.642	2558257	20.109 ug/ml
15) T Chrysene (C27.41)	14.684	2654097	20.044 ug/ml
16) T benzo[b]fluoranthene ...	16.184	2661676	20.295 ug/mlm
17) T Bnezo[k]fluoranthene ...	16.217	2649303	20.430 ug/ml
18) T Benzo[a]pyrene (C31.34)	16.560	2565870	19.990 ug/ml
19) T Indeno[1,2,3-cd]pyren...	17.918	2466204	19.635 ug/mlm
20) T Dibenz[a,h]anthracene...	17.958	2500914	20.448 ug/ml
21) T Benzo[g,h,i]perylene ...	18.169	2547428	20.079 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD022823AR\
Data File : FD045189.D
Signal(s) : FID2B.ch
Acq On : 28 Feb 2023 22:07
Operator : YP/AJ
Sample : 20 PPM AROMATIC HC STD
Misc :
ALS Vial : 52 Sample Multiplier: 1

Instrument :
FID_D
ClientSampleId :
20 PPM AROMATIC HC STD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/01/2023
Supervised By :Ankita Jodhani 03/01/2023

Integration File: autoint1.e
Quant Time: Mar 01 04:21:16 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aromatic EPH 022123.M
Quant Title : GC Extractables
QLast Update : Wed Feb 22 04:55:52 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18µm

