

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD030822AL\
 Data File : FD041431.D
 Signal(s) : FID2B.CH
 Acq On : 08 Mar 2022 12:49
 Operator : UA/NP
 Sample : N1806-10RE
 Misc :
 ALS Vial : 55 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Mar 09 00:34:22 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aliphatic EPH 022322.M
 Quant Title : GC Extractables
 QLast Update : Thu Feb 24 01:21:33 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.237	3149261	18.155 ug/ml
Spiked Amount 50.000		Recovery =	36.31%
12) S 1-chlorooctadecane (S...	12.678	2545734	18.804 ug/ml
Spiked Amount 50.000		Recovery =	37.61%
Target Compounds			
10) T n-Eicosane (C20)	12.270	113509	0.731 ug/ml
14) T n-Tetracosane (C24)	14.481f	96671	0.616 ug/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD030822AL\
Data File : FD041431.D
Signal(s) : FID2B.CH
Acq On : 08 Mar 2022 12:49
Operator : UA/NP
Sample : N1806-10RE
Misc :
ALS Vial : 55 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Mar 09 00:34:22 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aliphatic EPH 022322.M
Quant Title : GC Extractables
QLast Update : Thu Feb 24 01:21:33 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped


Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.237 min
Delta R.T.: -0.009 min
Response: 3149261
Conc: 18.16 ug/ml




#10 n-Eicosane (C20)

R.T.: 12.270 min
Delta R.T.: 0.022 min
Response: 113509
Conc: 0.73 ug/ml



#12 1-chlorooctadecane (SURR)

R.T.: 12.678 min
Delta R.T.: -0.008 min
Response: 2545734
Conc: 18.80 ug/ml



#14 n-Tetracosane (C24)

R.T.: 14.481 min
Delta R.T.: -0.054 min
Response: 96671
Conc: 0.62 ug/ml