

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD033021AR\
 Data File : FD036651.D
 Signal(s) : FID2B.CH
 Acq On : 30 Mar 2021 17:32
 Operator : DD\AJ
 Sample : M1858-07
 Misc :
 ALS Vial : 61 Sample Multiplier: 1

Instrument :
 FID_D
 ClientSampleId :
 B2B

Manual Integrations
 APPROVED

Ankita
 4/1/2021 8:45:03 AM

Integration File: autoint1.e
 Quant Time: Mar 31 01:07:39 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aromatic EPH 032321.M
 Quant Title : GC Extractables
 QLast Update : Wed Mar 24 11:18:39 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18µm

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
4) S 2-Bromonaphthalene (S...	7.594	10286499	73.139 ug/ml
Spiked Amount 50.000		Recovery =	146.28%
6) S 2-Fluorobiphenyl (SURR)	8.454	6671868	71.077 ug/ml
Spiked Amount 50.000 Range 0 - 131		Recovery =	142.15%#
11) S ortho-Terphenyl (SURR)	11.490	4956379	32.452 ug/mlm
Spiked Amount 50.000		Recovery =	64.90%

Target Compounds

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD033021AR\
Data File : FD036651.D
Signal(s) : FID2B.CH
Acq On : 30 Mar 2021 17:32
Operator : DD\AJ
Sample : M1858-07
Misc :
ALS Vial : 61 Sample Multiplier: 1


Instrument :
FID_D
ClientSampleId :
B2B


Manual Integrations
APPROVED

Ankita
4/1/2021 8:45:03 AM

Integration File: autoint1.e
Quant Time: Mar 31 01:07:39 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aromatic EPH 032321.M
Quant Title : GC Extractables
QLast Update : Wed Mar 24 11:18:39 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18µm






#4 2-Bromonaphthalene (SURR)

R.T.: 7.594 min
Delta R.T.: 0.000 min
Response: 10286499
Conc: 73.14 ug/ml

Instrument :
FID_D
ClientSampleId :
B2B


Manual Integrations
APPROVED

Ankita
4/1/2021 8:45:03 AM



#6 2-Fluorobiphenyl (SURR)

R.T.: 8.454 min
Delta R.T.: 0.001 min
Response: 6671868
Conc: 71.08 ug/ml



#11 ortho-Terphenyl (SURR)

R.T.: 11.490 min
Delta R.T.: -0.004 min
Response: 4956379
Conc: 32.45 ug/ml m