

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD040921AR\
 Data File : FD036850.D
 Signal(s) : FID2B.CH
 Acq On : 10 Apr 2021 12:34
 Operator : DD\AJ
 Sample : 20 PPM AROMATIC HC STD
 Misc :
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 FID_D
 LabSampleId :
 20 PPM AROMATIC HC STD

Integration File: sample.E
 Quant Time: Apr 12 01:12:44 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aromatic EPH 032321.M
 Quant Title : GC Extractables
 QLast Update : Wed Mar 24 11:18:39 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 T	1,2,3-Trimethylbenzene (C10	146.583	141.580 E3	3.4	96	0.00
2 T	Naphthalene (C11.7)	157.162	146.833 E3	6.6	93	0.00
3 T	2-Methylnaphthalene (C12.89	157.165	144.381 E3	8.1	91	0.00
4 S	2-Bromonaphthalene (SURR)	140.643	128.983 E3	8.3	91	0.00
5 T	Acenaphthylene (C15.06)	149.841	134.529 E3	10.2	89	0.00
6 S	2-Fluorobiphenyl (SURR)	93.869	83.415 E3	11.1	88	0.00
7 T	Acenaphthene (C15.5)	155.235	139.342 E3	10.2	89	0.00
8 T	Fluorene (C16.55)	150.558	132.515 E3	12.0	87	0.00
9 T	Phenanthrene (C19.36)	146.571	125.364 E3	14.5	84	0.00
10 T	Anthracene (C19.43)	144.768	114.310 E3	21.0	78	0.00
11 S	ortho-Terphenyl (SURR)	152.730	132.744 E3	13.1	86	0.00
12 T	Fluoranthene (C21.85)	140.396	117.440 E3	16.4	83	0.00
13 T	Pyrene (C20.8)	137.202	114.908 E3	16.2	83	0.00
14 T	Benzo[a]anthracene (C26.37)	118.758	107.091 E3	9.8	90	0.00
15 T	Chrysene (C27.41)	120.933	111.584 E3	7.7	91	0.00
16 T	benzo[b]fluoranthene (C30.4	109.732	111.775 E3	-1.9	102	0.00
17 T	Bnezo[k]fluoranthene (C30.1	106.298	108.625 E3	-2.2	102	0.00
18 T	Benzo[a]pyrene (C31.34)	99.662	104.146 E3	-4.5	104	0.00
19 T	Indeno[1,2,3-cd]pyrene (C35	100.518	110.014 E3	-9.4	110	0.00
20 T	Dibenz[a,h]anthracene (C30.	100.833	112.668 E3	-11.7	110	0.00
21 T	Benzo[g,h,i]perylene (C34.0	101.321	111.912 E3	-10.5	110	0.00

Evaluate Continuing Calibration Report - Not Found

(#) = Out of Range

SPCC's out = 0 CCC's out = 0