

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD040924AR\
 Data File : FD047506.D
 Signal(s) : FID2B.ch
 Acq On : 09 Apr 2024 08:14
 Operator : YP/AJ
 Sample : 20 PPM AROMATIC HC STD
 Misc :
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 FID_D
 LabSampleId :
 20 PPM AROMATIC HC STD

Integration File: autoint1.e
 Quant Time: Apr 09 12:59:52 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aromatic EPH 040224.M
 Quant Title : GC Extractables
 QLast Update : Mon Apr 01 16:48:10 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 T	1,2,3-Trimethylbenzene (C10	158.015	164.179 E3	-3.9	105	0.00
2 T	Naphthalene (C11.7)	163.641	171.509 E3	-4.8	106	0.00
3 T	2-Methylnaphthalene (C12.89	164.483	172.429 E3	-4.8	106	0.00
4 S	2-Bromonaphthalene (SURR)	149.501	156.918 E3	-5.0	106	0.00
5 T	Acenaphthylene (C15.06)	150.940	156.659 E3	-3.8	105	0.00
6 S	2-Fluorobiphenyl (SURR)	97.708	103.914 E3	-6.4	107	0.00
7 T	Acenaphthene (C15.5)	160.163	168.189 E3	-5.0	106	0.00
8 T	Flouorene (C16.55)	154.701	162.733 E3	-5.2	106	0.00
9 T	Phenanthrene (C19.36)	148.781	158.205 E3	-6.3	106	0.00
10 T	Anthracene (C19.43)	136.007	128.354 E3	5.6	95	0.00
11 S	ortho-Terphenyl (SURR)	171.425	181.238 E3	-5.7	107	0.00
12 T	Fluoranthene (C21.85)	147.830	157.552 E3	-6.6	108	0.00
13 T	Pyrene (C20.8)	144.321	153.192 E3	-6.1	108	0.00
14 T	Benzo[a]anthracene (C26.37)	131.513	139.739 E3	-6.3	107	0.00
15 T	Chrysene (C27.41)	141.010	151.430 E3	-7.4	109	0.00
16 T	benzo[b]fluoranthene (C30.4	139.513	148.478 E3	-6.4	108	0.00
17 T	Bnezo[k]fluoranthene (C30.1	139.089	149.314 E3	-7.4	109	0.00
18 T	Benzo[a]pyrene (C31.34)	137.303	144.292 E3	-5.1	106	0.00
19 T	Indeno[1,2,3-cd]pyrene (C35	134.890	141.382 E3	-4.8	106	0.00
20 T	Dibenz[a,h]anthracene (C30.	134.398	150.213 E3	-11.8	112	0.00
21 T	Benzo[g,h,i]perylene (C34.0	138.790	145.651 E3	-4.9	106	0.00

Evaluate Continuing Calibration Report - Not Founds

(#) = Out of Range

SPCC's out = 0 CCC's out = 0