

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD042121AR\
 Data File : FD037073.D
 Signal(s) : FID2B.CH
 Acq On : 22 Apr 2021 10:55
 Operator : DD\AJ
 Sample : M2090-13
 Misc :
 ALS Vial : 80 Sample Multiplier: 1

Instrument :
 FID_D
 ClientSampleId :
 SEDGWICK-COMP-3

Manual Integrations
 APPROVED

Ankita
 4/22/2021 4:48:53 PM

Integration File: autoint1.e
 Quant Time: Apr 22 13:22:33 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aromatic EPH 042121.M
 Quant Title : GC Extractables
 QLast Update : Thu Apr 22 07:01:17 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18µm

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
4) S 2-Bromonaphthalene (S...	7.576	5355550	45.295 ug/ml
Spiked Amount 50.000		Recovery =	90.59%
6) S 2-Fluorobiphenyl (SURR)	8.436	3245937	42.066 ug/ml
Spiked Amount 50.000 Range 0 - 131		Recovery =	84.13%
11) S ortho-Terphenyl (SURR)	11.475	2364530	19.143 ug/mlm
Spiked Amount 50.000		Recovery =	38.29%

Target Compounds

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD042121AR\
Data File : FD037073.D
Signal(s) : FID2B.CH
Acq On : 22 Apr 2021 10:55
Operator : DD\AJ
Sample : M2090-13
Misc :
ALS Vial : 80 Sample Multiplier: 1


Instrument :
FID_D
ClientSampleId :
SEDGWICK-COMP-3


Manual Integrations
APPROVED

Ankita
4/22/2021 4:48:53 PM

Integration File: autoint1.e
Quant Time: Apr 22 13:22:33 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aromatic EPH 042121.M
Quant Title : GC Extractables
QLast Update : Thu Apr 22 07:01:17 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18µm






#4 2-Bromonaphthalene (SURR)

R.T.: 7.576 min
Delta R.T.: 0.002 min
Response: 5355550
Conc: 45.30 ug/ml

Instrument :
FID_D
ClientSampleId :
SEDGWICK-COMP-3


Manual Integrations
APPROVED

Ankita
4/22/2021 4:48:53 PM



#6 2-Fluorobiphenyl (SURR)

R.T.: 8.436 min
Delta R.T.: 0.002 min
Response: 3245937
Conc: 42.07 ug/ml



#11 ortho-Terphenyl (SURR)

R.T.: 11.475 min
Delta R.T.: 0.000 min
Response: 2364530
Conc: 19.14 ug/ml m