

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC042622AL\
 Data File : FC059460.D
 Signal(s) : FID1A.CH
 Acq On : 27 Apr 2022 4:05
 Operator : UA/NP
 Sample : N2349-13DL2 200X
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 FID_D
 ClientSampleId :

Integration File: autoint1.e
 Quant Time: Apr 27 05:47:20 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 042122.M
 Quant Title : GC Extractables
 QLast Update : Thu Apr 21 17:30:29 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
12) S 1-chlorooctadecane (S...	12.544	70132	0.537 ug/ml
Spiked Amount 50.000		Recovery =	1.07%
Target Compounds			
11) T n-Heneicosane (C21)	12.713	41966	0.276 ug/ml
13) T n-Docosane (C22)	13.299	187510	1.224 ug/ml
14) T n-Tetracosane (C24)	14.395	251979	1.627 ug/ml
16) T n-Octacosane (C28)	16.336	191053	1.234 ug/ml
17) T n-Tricontane (C30)	17.171f	210337	1.385 ug/ml
19) T n-Tetratriacontane (C34)	18.867	246548	1.951 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC042622AL\
Data File : FC059460.D
Signal(s) : FID1A.CH
Acq On : 27 Apr 2022 4:05
Operator : UA/NP
Sample : N2349-13DL2 200X
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
FID_D
ClientSampleId :

Integration File: autoint1.e
Quant Time: Apr 27 05:47:20 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 042122.M
Quant Title : GC Extractables
QLast Update : Thu Apr 21 17:30:29 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)


R.T.: 0.000 min
Exp R.T.: 11.097 min
Response: 0
Conc: N.D.

Instrument :
FID_D
ClientSampleId :




#11 n-Heneicosane (C21)

R.T.: 12.713 min
Delta R.T.: -0.001 min
Response: 41966
Conc: 0.28 ug/ml




#12 1-chlorooctadecane (SURR)

R.T.: 12.544 min
Delta R.T.: 0.000 min
Response: 70132
Conc: 0.54 ug/ml



#13 n-Docosane (C22)


R.T.: 13.299 min
Delta R.T.: 0.001 min
Response: 187510
Conc: 1.22 ug/ml



#14 n-Tetracosane (C24)


R.T.: 14.395 min
Delta R.T.: 0.001 min
Response: 251979
Conc: 1.63 ug/ml

Instrument :
FID_D
ClientSampleId :




#16 n-Octacosane (C28)

R.T.: 16.336 min
Delta R.T.: -0.020 min
Response: 191053
Conc: 1.23 ug/ml



#17 n-Tricontane (C30)

R.T.: 17.171 min
Delta R.T.: -0.068 min
Response: 210337
Conc: 1.38 ug/ml



#19 n-Tetratriacontane (C34)

R.T.: 18.867 min
Delta R.T.: 0.017 min
Response: 246548
Conc: 1.95 ug/ml