

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD042722AR\
 Data File : FD042134.D
 Signal(s) : FID2B.CH
 Acq On : 27 Apr 2022 12:31
 Operator : UA/NP
 Sample : 20 PPM AROMATIC HC STD
 Misc :
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 FID_D
 ClientSampleId :
 20 PPM AROMATIC HC STD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/28/2022
 Supervised By :mohammad ahmed 04/28/2022

Integration File: sample.E
 Quant Time: Apr 28 03:31:34 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aromatic EPH 042122.M
 Quant Title : GC Extractables
 QLast Update : Thu Apr 21 17:32:22 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18µm

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
4) S 2-Bromonaphthalene (S...	7.328	2881990	20.561 ug/ml
Spiked Amount 50.000		Recovery =	41.12%
6) S 2-Fluorobiphenyl (SURR)	8.179	1622340	22.769 ug/ml
Spiked Amount 50.000 Range 0 - 131		Recovery =	45.54%
11) S ortho-Terphenyl (SURR)	11.204	3272549	18.880 ug/ml
Spiked Amount 50.000		Recovery =	37.76%
Target Compounds			
1) T 1,2,3-Trimethylbenzen...	4.151	3162527	20.314 ug/ml
2) T Naphthalene (C11.7)	5.672	3299831	20.918 ug/ml
3) T 2-Methylnaphthalene (...)	6.717	3158118	21.840 ug/ml
5) T Acenaphthylene (C15.06)	7.978	3239795	19.659 ug/ml
7) T Acenaphthene (C15.5)	8.271	3810001	20.294 ug/ml
8) T Flouorene (C16.55)	9.050	3038350	19.948 ug/ml
9) T Phenanthrene (C19.36)	10.437	2893028	19.483 ug/ml
10) T Anthracene (C19.43)	10.512	3040061	18.159 ug/ml
12) T Fluoranthene (C21.85)	12.238	2944196	18.770 ug/ml
13) T Pyrene (C20.8)	12.530	2866742	17.576 ug/ml
14) T Benzo[a]anthracene (C...	14.402	2393340	20.008 ug/ml
15) T Chrysene (C27.41)	14.442	3279390	17.875 ug/ml
16) T benzo[b]fluoranthene ...	15.949	2460746	18.742 ug/ml
17) T Bnezo[k]fluoranthene ...	15.980	3010982	17.416 ug/ml
18) T Benzo[a]pyrene (C31.34)	16.329	2482548	16.826 ug/ml
19) T Indeno[1,2,3-cd]pyren...	17.692	2119877	21.239 ug/mlm
20) T Dibenz[a,h]anthracene...	17.731	2940371	18.734 ug/ml
21) T Benzo[g,h,i]perylene ...	17.941	2469194	17.829 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD042722AR\
Data File : FD042134.D
Signal(s) : FID2B.CH
Acq On : 27 Apr 2022 12:31
Operator : UA/NP
Sample : 20 PPM AROMATIC HC STD
Misc :
ALS Vial : 52 Sample Multiplier: 1

Instrument :
FID_D
ClientSampleId :
20 PPM AROMATIC HC STD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/28/2022
Supervised By :mohammad ahmed 04/28/2022

Integration File: sample.E
Quant Time: Apr 28 03:31:34 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aromatic EPH 042122.M
Quant Title : GC Extractables
QLast Update : Thu Apr 21 17:32:22 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18µm

