

Data Path : P:\HPCHEM1\FID D\Data\FD051618AL\
 Data File : FD023731.D
 Signal(s) : FID2B.CH
 Acq On : 17 May 2018 9:47
 Operator : UA/AJ
 Sample : 20 PPM ALIPHATIC HC STD
 Misc :
 ALS Vial : 56 Sample Multiplier: 1

Instrument :
 FID_D
 LabSampleId :
 20 PPM ALIPHATIC HC STD

Integration File: autoint1.e
 Quant Time: May 17 13:15:22 2018
 Quant Method : P:\HPCHEM1\FID_D\methods\Alaphatic EPH 051518.M
 Quant Title : GC Extractables
 QLast Update : Wed May 16 11:11:38 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvaRF	CCRF	%Dev	Area%	Dev(Min)
1 T	n-Nonane (C9)	103.054	97.329 E3	5.6	97	0.00
2 T	n-Decane (C10)	102.491	96.678 E3	5.7	97	0.00
3 T	A-Naphthalene (C11.7)	111.697	105.503 E3	5.5	97	0.00
4 T	n-Dodecane (C12)	101.142	94.810 E3	6.3	96	0.00
5 T	A-2-methylnaphthalene (C12)	110.980	104.947 E3	5.4	97	0.00
6 T	n-Tetradecane (C14)	100.352	94.019 E3	6.3	96	0.00
7 T	n-Hexadecane (C16)	99.385	93.332 E3	6.1	96	0.00
8 T	n-Octadecane (C18)	97.322	91.889 E3	5.6	97	0.00
9 S	ortho-Terphenyl (SURR)	105.630	100.991 E3	4.4	98	0.00
10 T	n-Eicosane (C20)	99.441	94.723 E3	4.7	98	0.00
11 T	n-Heneicosane (C21)	99.545	94.552 E3	5.0	98	0.00
12 S	1-chlorooctadecane (SURR)	83.966	80.635 E3	4.0	99	0.00
13 T	n-Docosane (C22)	100.389	96.379 E3	4.0	99	0.00
14 T	n-Tetracosane (C24)	100.840	98.984 E3	1.8	101	0.00
15 T	n-Hexacosane (C26)	100.688	100.292 E3	0.4	103	0.00
16 T	n-Octacosane (C28)	102.088	102.953 E3	-0.8	104	0.00
17 T	n-Tricontane (C30)	99.074	99.524 E3	-0.5	104	0.00
18 T	n-Dotriacontane (C32)	93.524	94.683 E3	-1.2	105	0.00
19 T	n-Tetratriacontane (C34)	78.305	77.366 E3	1.2	100	0.00
20 T	n-Hexatriacontane (C36)	53.232	54.057 E3	-1.5	99	0.00
21 T	n-Octatriacontane (C38)	34.769	41.140 E3	-18.3	106	0.00
22 T	n-Tetracontane (C40)	24.588	22.378 E3	9.0	76	0.00

Evaluate Continuing Calibration Report - Not Found

(#) = Out of Range

SPCC's out = 0 CCC's out = 0