

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD051923AR\
 Data File : FD045707.D
 Signal(s) : FID2B.ch
 Acq On : 20 May 2023 12:31
 Operator : YP/AJ
 Sample : 02837-01DL2 50X
 Misc :
 ALS Vial : 59 Sample Multiplier: 1

Instrument :
 FID_D
 ClientSampleId :
 WASTE-1DL2

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/22/2023
 Supervised By :Ankita Jodhani 05/22/2023

Integration File: autoint1.e
 Quant Time: May 21 04:46:36 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aromatic EPH 050523.M
 Quant Title : GC Extractables
 QLast Update : Fri May 05 18:09:33 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18µm

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
4) S 2-Bromonaphthalene (S...	7.541	317719	1.863 ug/mlm
Spiked Amount 50.000		Recovery =	3.73%
6) S 2-Fluorobiphenyl (SURR)	8.394	342569	2.980 ug/mlm
Spiked Amount 50.000 Range 0 - 131		Recovery =	5.96%
11) S ortho-Terphenyl (SURR)	11.433	525773	2.490 ug/mlm
Spiked Amount 50.000		Recovery =	4.98%

Target Compounds

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD051923AR\
Data File : FD045707.D
Signal(s) : FID2B.ch
Acq On : 20 May 2023 12:31
Operator : YP/AJ
Sample : 02837-01DL2 50X
Misc :
ALS Vial : 59 Sample Multiplier: 1

Instrument :

FID_D

ClientSampleId :

WASTE-1DL2

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 05/22/2023

Supervised By :Ankita Jodhani 05/22/2023

Integration File: autoint1.e

Quant Time: May 21 04:46:36 2023

Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aromatic EPH 050523.M

Quant Title : GC Extractables

QLast Update : Fri May 05 18:09:33 2023


Response via : Initial Calibration


Integrator: ChemStation

Volume Inj. : 1 µl

Signal Phase : Rxi-1ms

Signal Info : 20M x 0.18mm x 0.18µm






#4 2-Bromonaphthalene (SURR)

R.T.: 7.541 min
Delta R.T.: 0.000 min
Response: 317719
Conc: 1.86 ug/ml

Instrument :
FID_D
ClientSampleId :
WASTE-1DL2


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/22/2023
Supervised By :Ankita Jodhani 05/22/2023



#6 2-Fluorobiphenyl (SURR)

R.T.: 8.394 min
Delta R.T.: -0.003 min
Response: 342569
Conc: 2.98 ug/ml m



#11 ortho-Terphenyl (SURR)

R.T.: 11.433 min
Delta R.T.: -0.003 min
Response: 525773
Conc: 2.49 ug/ml m