

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD053122AL\
 Data File : FD042702.D
 Signal(s) : FID2B.CH
 Acq On : 01 Jun 2022 1:48
 Operator : UA/NP
 Sample : N3083-11
 Misc :
 ALS Vial : 62 Sample Multiplier: 1

Instrument :
 FID_D
ClientSampleId :
 D-17-5-SW-SB-38(9.5-10)

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/01/2022
 Supervised By :Ankita Jodhani 06/01/2022

Integration File: sample.E
 Quant Time: Jun 01 03:09:38 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aliphatic EPH 052722.M
 Quant Title : GC Extractables
 QLast Update : Sat May 28 07:55:57 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

| Compound | R.T. | Response | Conc Units |
|--------------------------------|--------|----------|---------------|
| ----- | | | |
| System Monitoring Compounds | | | |
| 9) S ortho-Terphenyl (SURR) | 11.620 | 11738533 | 70.679 ug/mlm |
| Spiked Amount | 50.000 | Recovery | = 141.36% |
| 12) S 1-chlorooctadecane (S... | 13.019 | 2298678 | 17.601 ug/mlm |
| Spiked Amount | 50.000 | Recovery | = 35.20% |

Target Compounds

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD053122AL\
 Data File : FD042702.D
 Signal(s) : FID2B.CH
 Acq On : 01 Jun 2022 1:48
 Operator : UA/NP
 Sample : N3083-11
 Misc :
 ALS Vial : 62 Sample Multiplier: 1


Instrument :
 FID_D
ClientSampleId :
 D-17-5-SW-SB-38(9.5-10)


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/01/2022
 Supervised By :Ankita Jodhani 06/01/2022

Integration File: sample.E
 Quant Time: Jun 01 03:09:38 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aliphatic EPH 052722.M
 Quant Title : GC Extractables
 QLast Update : Sat May 28 07:55:57 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.620 min
 Delta R.T.: 0.039 min
 Response: 11738533
 Conc: 70.68 ug/ml

Instrument : FID_D
 ClientSampleId : D-17-5-SW-SB-38(9.5-10)

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/01/2022
 Supervised By :Ankita Jodhani 06/01/2022



#12 1-chlorooctadecane (SURR)

R.T.: 13.019 min
 Delta R.T.: 0.005 min
 Response: 2298678
 Conc: 17.60 ug/ml m