

rteres

Instrument :
FID_D

LabSampleId :
20 PPM ALIPHATIC HC STD

Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 06/03/2022
Supervised By :Ankita Jodhani 06/03/2022

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD06022
Data File : FD042753.D
Signal(s) : FID2B.CH
Acq On : 03 Jun 2022 2:37
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aliphatic EPH 052722.M
Title : GC Extractables

Signal : FID2B.CH

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.352	3.315	3.391	BB	333826	2755352	85.76%	4.739%
2	4.420	4.381	4.455	BB	304139	2743125	85.38%	4.718%
3	6.003	5.959	6.034	BB	343158	3158458	98.31%	5.433%
4	6.436	6.397	6.478	BB	291564	2819659	87.76%	4.850%
5	7.057	7.014	7.099	BB	337773	3212767	100.00%	5.526%
6	8.228	8.182	8.270	BB	272106	2852733	88.79%	4.907%
7	9.828	9.782	9.868	BB	264841	2893058	90.05%	4.976%
8	11.268	11.223	11.308	BB	253325	2914160	90.71%	5.013%
9	11.584	11.542	11.628	BB	293616	3159019	98.33%	5.434%
10	12.575	12.524	12.615	BB	244741	2861280	89.06%	4.922%
11	13.018	12.974	13.058	BB	212949	2477418	77.11%	4.261%
12	13.186	13.140	13.234	BB	238881	2854708	88.86%	4.910%
13	13.772	13.734	13.807	BB	228149	2821997	87.84%	4.854%
14	14.872	14.778	14.922	BB	218921	2753727	85.71%	4.737%
15	15.890	15.844	15.949	BB	199881	2620379	81.56%	4.507%
16	16.839	16.787	16.878	BB	183454	2493102	77.60%	4.288%
17	17.724	17.674	17.769	BB	167526	2372741	73.85%	4.081%
18	18.555	18.499	18.619	BB	151090	2270691	70.68%	3.906%
19	19.336	19.281	19.382	BB	140608	2065497	64.29%	3.553%
20	20.074	20.031	20.107	M	135639	1971619	61.37%	3.391%
21	20.807	20.760	20.848	M	110509	1909581	59.44%	3.285%
22	21.709	21.621	21.834	BV	84720	2155583	67.09%	3.708%
Sum of corrected areas:							58136651	

Aliphatic EPH 052722.M Fri Jun 03 11:07:29 2022