

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD060222AL\
 Data File : FD042739.D
 Signal(s) : FID2B.CH
 Acq On : 02 Jun 2022 15:20
 Operator : UA/NP
 Sample : N3084-17
 Misc :
 ALS Vial : 60 Sample Multiplier: 1

Instrument :
 FID_D
 ClientSampleId :
 D-17-8-SW-SB-2(9.5-10)

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/03/2022
 Supervised By :Ankita Jodhani 06/03/2022

Integration File: sample.E
 Quant Time: Jun 02 15:49:35 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aliphatic EPH 052722.M
 Quant Title : GC Extractables
 QLast Update : Sat May 28 07:55:57 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.593	6708438	40.392 ug/mlm
Spiked Amount	50.000	Recovery	= 80.78%
12) S 1-chlorooctadecane (S...	13.031	4752644	36.391 ug/mlm
Spiked Amount	50.000	Recovery	= 72.78%

Target Compounds

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD060222AL\
Data File : FD042739.D
Signal(s) : FID2B.CH
Acq On : 02 Jun 2022 15:20
Operator : UA/NP
Sample : N3084-17
Misc :
ALS Vial : 60 Sample Multiplier: 1


Instrument :
FID_D
ClientSampleId :
D-17-8-SW-SB-2(9.5-10)


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/03/2022
Supervised By :Ankita Jodhani 06/03/2022

Integration File: sample.E
Quant Time: Jun 02 15:49:35 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aliphatic EPH 052722.M
Quant Title : GC Extractables
QLast Update : Sat May 28 07:55:57 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um





#9 ortho-Terphenyl (SURR)

R.T.: 11.593 min
Delta R.T.: 0.012 min
Response: 6708438
Conc: 40.39 ug/ml

Instrument :

FID_D

ClientSampleId :


D-17-8-SW-SB-2(9.5-10)

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 06/03/2022

Supervised By :Ankita Jodhani 06/03/2022



#12 1-chlorooctadecane (SURR)

R.T.: 13.031 min
Delta R.T.: 0.017 min
Response: 4752644
Conc: 36.39 ug/ml m