

Instrument :
FID_D
LabSampleId :
5 PPM ALIPHATIC HC STD5

rteres

Area Percent Report

Manual Integrations
APPROVED

Ankita
6/11/2021 2:29:06 PM

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD061021AL\
Data File : FD037853.D
Signal(s) : FID2B.CH
Acq On : 10 Jun 2021 17:17
Sample : 5 PPM ALIPHATIC HC STD5
Misc :
ALS Vial : 60 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Alaphatic EPH 061021.M
Title : GC Extractables

Signal : FID2B.CH

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.249	3.223	3.291	BB	76225	616460	85.56%	4.478%
2	4.307	4.282	4.341	BB	68918	615782	85.47%	4.473%
3	5.879	5.851	5.922	BB	73214	667654	92.67%	4.850%
4	6.313	6.284	6.348	BB	65367	621167	86.21%	4.512%
5	6.930	6.898	6.972	BB	67830	664011	92.16%	4.824%
6	8.100	8.068	8.129	BB	63083	631261	87.62%	4.586%
7	9.694	9.657	9.735	BB	60342	647339	89.85%	4.703%
8	11.130	11.092	11.167	BB	59614	670989	93.13%	4.874%
9	11.440	11.405	11.476	BB	67884	720490	100.00%	5.234%
10	12.434	12.393	12.479	BB	60043	675094	93.70%	4.904%
11	12.875	12.835	12.910	BB	49006	568399	78.89%	4.129%
12	13.043	13.000	13.083	BB	58540	684716	95.03%	4.974%
13	13.627	13.587	13.668	BB	56239	675634	93.77%	4.908%
14	14.724	14.682	14.763	BB	55135	679181	94.27%	4.934%
15	15.741	15.698	15.779	BB	52969	662728	91.98%	4.814%
16	16.686	16.641	16.721	BB	51390	653373	90.68%	4.746%
17	17.570	17.526	17.615	BB	50207	650165	90.24%	4.723%
18	18.398	18.349	18.443	BB	44761	621625	86.28%	4.516%
19	19.177	19.131	19.218	BB	42160	576346	79.99%	4.187%
20	19.913	19.868	19.946	M	37171	519699	72.13%	3.775%
21	20.622	20.569	20.669	BB	29928	466496	64.75%	3.389%
22	21.459	21.406	21.519	BB	21751	477010	66.21%	3.465%
Sum of corrected areas:						13765618		

Alaphatic EPH 061021.M Fri Jun 11 12:19:48 2021