

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_D\Data\FD061122AL\  
 Data File : FD042872.D  
 Signal(s) : FID2B.CH  
 Acq On : 11 Jun 2022 1:19  
 Operator : UA/NP  
 Sample : N3299-06  
 Misc :  
 ALS Vial : 62 Sample Multiplier: 1

Instrument :  
 FID\_D  
 ClientSampleId :  
 SC-18

Integration File: sample.E  
 Quant Time: Jun 11 02:45:13 2022  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_D\methods\Aliphatic EPH 052722.M  
 Quant Title : GC Extractables  
 QLast Update : Sat May 28 07:55:57 2022  
 Response via : Initial Calibration  
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul  
 Signal Phase : Rxi-1ms  
 Signal Info : 20M x 0.18mm x 0.18um

| Compound                       | R.T.   | Response   | Conc Units   |
|--------------------------------|--------|------------|--------------|
| -----                          |        |            |              |
| System Monitoring Compounds    |        |            |              |
| 9) S ortho-Terphenyl (SURR)    | 11.581 | 10514548   | 63.309 ug/ml |
| Spiked Amount 50.000           |        | Recovery = | 126.62%      |
| 12) S 1-chlorooctadecane (S... | 13.015 | 8615228    | 65.966 ug/ml |
| Spiked Amount 50.000           |        | Recovery = | 131.93%      |

Target Compounds

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_D\Data\FD061122AL\  
Data File : FD042872.D  
Signal(s) : FID2B.CH  
Acq On : 11 Jun 2022 1:19  
Operator : UA/NP  
Sample : N3299-06  
Misc :  
ALS Vial : 62 Sample Multiplier: 1

Instrument :  
FID\_D  
ClientSampleId :  
SC-18

Integration File: sample.E  
Quant Time: Jun 11 02:45:13 2022  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_D\methods\Aliphatic EPH 052722.M  
Quant Title : GC Extractables  
QLast Update : Sat May 28 07:55:57 2022  
Response via : Initial Calibration  
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul  
Signal Phase : Rxi-1ms  
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.581 min  
Delta R.T.: 0.000 min  
Response: 10514548  
Conc: 63.31 ug/ml

Instrument :  
FID\_D  
ClientSampleId :  
SC-18



#12 1-chlorooctadecane (SURR)

R.T.: 13.015 min  
Delta R.T.: 0.001 min  
Response: 8615228  
Conc: 65.97 ug/ml