

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD061822AR\
 Data File : FD042950.D
 Signal(s) : FID2B.CH
 Acq On : 18 Jun 2022 15:23
 Operator : UA/NP
 Sample : I.BLK
 Misc :
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 FID_D
 ClientSampleId :

Integration File: autoint1.e
 Quant Time: Jun 20 02:44:20 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aromatic EPH 052722.M
 Quant Title : GC Extractables
 QLast Update : Wed Jun 01 10:34:31 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18µm

Compound	R.T.	Response	Conc Units
----------	------	----------	------------

System Monitoring Compounds

Target Compounds

14) T	Benzo[a]anthracene (C...	14.814	132351	0.993 ug/ml
15) T	Chrysene (C27.41)	14.814	132351	1.002 ug/ml
16) T	benzo[b]fluoranthene ...	16.354	96594	0.725 ug/ml
17) T	Bnezo[k]fluoranthene ...	16.354	96594	0.780 ug/ml
19) T	Indeno[1,2,3-cd]pyren...	18.126	98865	0.887 ug/ml
20) T	Dibenz[a,h]anthracene...	18.126	98865	0.889 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD061822AR\
Data File : FD042950.D
Signal(s) : FID2B.CH
Acq On : 18 Jun 2022 15:23
Operator : UA/NP
Sample : I.BLK
Misc :
ALS Vial : 51 Sample Multiplier: 1

Instrument :
FID_D
ClientSampleId :

Integration File: autoint1.e
Quant Time: Jun 20 02:44:20 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aromatic EPH 052722.M
Quant Title : GC Extractables
QLast Update : Wed Jun 01 10:34:31 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18µm






#4 2-Bromonaphthalene (SURRE)


R.T.: 0.000 min
Exp R.T. : 7.667 min
Response: 0
Conc: N.D.

Instrument :
FID_D
ClientSampleId :




#6 2-Fluorobiphenyl (SURRE)

R.T.: 0.000 min
Exp R.T. : 8.529 min
Response: 0
Conc: N.D.




#11 ortho-Terphenyl (SURRE)

R.T.: 0.000 min
Exp R.T. : 11.576 min
Response: 0
Conc: N.D.



#14 Benzo[a]anthracene (C26.37)


R.T.: 14.814 min
Delta R.T.: 0.030 min
Response: 132351
Conc: 0.99 ug/ml



#15 Chrysene (C27.41)


R.T.: 14.814 min
Delta R.T.: -0.011 min
Response: 132351
Conc: 1.00 ug/ml

Instrument :
FID_D
ClientSampleId :




#16 benzo[b]fluoranthene (C30.41)

R.T.: 16.354 min
Delta R.T.: 0.022 min
Response: 96594
Conc: 0.73 ug/ml




#17 Bnezo[k]fluoranthene (C30.14)

R.T.: 16.354 min
Delta R.T.: -0.009 min
Response: 96594
Conc: 0.78 ug/ml



#19 Indeno[1,2,3-cd]pyrene (C35.01)

R.T.: 18.126 min
Delta R.T.: 0.055 min
Response: 98865
Conc: 0.89 ug/ml



#20 Dibenz[a,h]anthracene (C30.36)

R.T.: 18.126 min
Delta R.T.: 0.013 min
Response: 98865
Conc: 0.89 ug/ml

Instrument :
FID_D
ClientSampleId :