

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD070120AL\
 Data File : FD033850.D
 Signal(s) : FID2B.CH
 Acq On : 01 Jul 2020 18:37
 Operator : DD\AJ
 Sample : L3162-02D
 Misc :
 ALS Vial : 58 Sample Multiplier: 1

Instrument :
 FID_D
 ClientSampleId :
 GL-SP

Integration File: autoint1.e
 Quant Time: Jul 02 02:44:07 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Alaphatic EPH 061820.M
 Quant Title : GC Extractables
 QLast Update : Fri Jun 19 09:06:59 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.451	5358048	37.491 ug/ml
Spiked Amount 50.000		Recovery =	74.98%
12) S 1-chlorooctadecane (S...	12.881	4291210	36.496 ug/ml
Spiked Amount 50.000		Recovery =	72.99%
Target Compounds			
10) T n-Eicosane (C20)	12.504	295521	2.171 ug/ml
14) T n-Tetracosane (C24)	14.717	246631	1.804 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD070120AL\
Data File : FD033850.D
Signal(s) : FID2B.CH
Acq On : 01 Jul 2020 18:37
Operator : DD\AJ
Sample : L3162-02D
Misc :
ALS Vial : 58 Sample Multiplier: 1

Instrument :
FID_D
ClientSampleId :
GL-SP

Integration File: autoint1.e
Quant Time: Jul 02 02:44:07 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Alaphatic EPH 061820.M
Quant Title : GC Extractables
QLast Update : Fri Jun 19 09:06:59 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)


R.T.: 11.451 min
Delta R.T.: -0.016 min
Response: 5358048
Conc: 37.49 ug/ml

Instrument :
FID_D
ClientSampleId :
GL-SP




#10 n-Eicosane (C20)

R.T.: 12.504 min
Delta R.T.: 0.050 min
Response: 295521
Conc: 2.17 ug/ml



#12 1-chlorooctadecane (SURR)

R.T.: 12.881 min
Delta R.T.: -0.014 min
Response: 4291210
Conc: 36.50 ug/ml



#14 n-Tetracosane (C24)

R.T.: 14.717 min
Delta R.T.: -0.024 min
Response: 246631
Conc: 1.80 ug/ml