

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD081623AR\
 Data File : FD046332.D
 Signal(s) : FID2B.ch
 Acq On : 16 Aug 2023 11:01
 Operator : YP/AJ
 Sample : 20 PPM AROMATIC HC STD
 Misc :
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 FID_D
 ClientSampleId :
 20 PPM AROMATIC HC STD

Integration File: autoint1.e
 Quant Time: Aug 16 23:46:55 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aromatic EPH 071923.M
 Quant Title : GC Extractables
 QLast Update : Thu Jul 20 03:07:14 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18µm

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
4) S 2-Bromonaphthalene (S...	7.499	2824024	18.892 ug/ml
Spiked Amount 50.000		Recovery =	37.78%
6) S 2-Fluorobiphenyl (SURR)	8.352	1866267	19.936 ug/ml
Spiked Amount 50.000 Range 0 - 131		Recovery =	39.87%
11) S ortho-Terphenyl (SURR)	11.389	3635298	20.816 ug/ml
Spiked Amount 50.000		Recovery =	41.63%
Target Compounds			
1) T 1,2,3-Trimethylbenzen...	4.312	3142072	17.246 ug/ml
2) T Naphthalene (C11.7)	5.839	3282744	18.175 ug/ml
3) T 2-Methylnaphthalene (...)	6.887	3286887	18.675 ug/ml
5) T Acenaphthylene (C15.06)	8.156	2983910	19.557 ug/ml
7) T Acenaphthene (C15.5)	8.451	3179651	19.500 ug/ml
8) T Fluorene (C16.55)	9.229	3099895	20.373 ug/ml
9) T Phenanthrene (C19.36)	10.618	3102816	21.548 ug/ml
10) T Anthracene (C19.43)	10.692	2790772	21.325 ug/ml
12) T Fluoranthene (C21.85)	12.420	3093147	22.136 ug/ml
13) T Pyrene (C20.8)	12.715	2980455	21.505 ug/ml
14) T Benzo[a]anthracene (C...	14.578	2681190	20.427 ug/ml
15) T Chrysene (C27.41)	14.620	2781009	19.723 ug/ml
16) T benzo[b]fluoranthene ...	16.120	2605136	19.048 ug/ml
17) T Bnezo[k]fluoranthene ...	16.151	2555070	18.641 ug/ml
18) T Benzo[a]pyrene (C31.34)	16.494	2437685	18.724 ug/ml
19) T Indeno[1,2,3-cd]pyren...	17.850	2160275	17.961 ug/ml
20) T Dibenz[a,h]anthracene...	17.892	2138592	16.721 ug/ml
21) T Benzo[g,h,i]perylene ...	18.101	2278759	17.962 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Instrument :
FID_D
ClientSampleId :
20 PPM AROMATIC HC STD

Volume Inj. : 1 µl
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18µm

