

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD090820AL\
 Data File : FD034688.D
 Signal(s) : FID2B.CH
 Acq On : 09 Sep 2020 1:55
 Operator : DD\AJ
 Sample : L3919-17D
 Misc :
 ALS Vial : 69 Sample Multiplier: 1

Instrument :
 FID_D
 ClientSampleId :

Integration File: autoint1.e
 Quant Time: Sep 09 06:19:17 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Alaphatic EPH 083120.M
 Quant Title : GC Extractables
 QLast Update : Tue Sep 01 01:55:21 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.597	4566476	36.957 ug/ml
Spiked Amount 50.000		Recovery =	73.91%
12) S 1-chlorooctadecane (S...	13.026	3614003	36.899 ug/ml
Spiked Amount 50.000		Recovery =	73.80%
Target Compounds			
10) T n-Eicosane (C20)	12.650f	113959	0.974 ug/ml
11) T n-Heneicosane (C21)	13.252f	66991	0.574 ug/ml
17) T n-Tricontane (C30)	17.653f	25173	0.232 ug/ml
18) T n-Dotriacontane (C32)	18.539	215840	2.142 ug/ml
19) T n-Tetratriacontane (C34)	19.388f	129175	1.368 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD090820AL\
Data File : FD034688.D
Signal(s) : FID2B.CH
Acq On : 09 Sep 2020 1:55
Operator : DD\AJ
Sample : L3919-17D
Misc :
ALS Vial : 69 Sample Multiplier: 1

Instrument :
FID_D
ClientSampleId :

Integration File: autoint1.e
Quant Time: Sep 09 06:19:17 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Alaphatic EPH 083120.M
Quant Title : GC Extractables
QLast Update : Tue Sep 01 01:55:21 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)


R.T.: 11.597 min
Delta R.T.: 0.003 min
Response: 4566476
Conc: 36.96 ug/ml

Instrument :
FID_D
ClientSampleId :




#10 n-Eicosane (C20)

R.T.: 12.650 min
Delta R.T.: 0.070 min
Response: 113959
Conc: 0.97 ug/ml




#11 n-Heneicosane (C21)

R.T.: 13.252 min
Delta R.T.: 0.062 min
Response: 66991
Conc: 0.57 ug/ml



#12 1-chlorooctadecane (SURR)


R.T.: 13.026 min
Delta R.T.: 0.002 min
Response: 3614003
Conc: 36.90 ug/ml



#17 n-Tricontane (C30)


R.T.: 17.653 min
Delta R.T.: -0.063 min
Response: 25173
Conc: 0.23 ug/ml

Instrument :
FID_D
ClientSampleId :



#18 n-Dotriacontane (C32)

R.T.: 18.539 min
Delta R.T.: -0.005 min
Response: 215840
Conc: 2.14 ug/ml



#19 n-Tetratriacontane (C34)

R.T.: 19.388 min
Delta R.T.: 0.064 min
Response: 129175
Conc: 1.37 ug/ml