

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_D\Data\FD092220AR\  
 Data File : FD034805.D  
 Signal(s) : FID2B.CH  
 Acq On : 22 Sep 2020 12:46  
 Operator : DD\AJ  
 Sample : SURROGATE TEST  
 Misc :  
 ALS Vial : 54 Sample Multiplier: 1

Instrument :  
 FID\_D  
 ClientSampleId :  
 SURROGATE TEST

Integration File: autoint1.e  
 Quant Time: Sep 23 02:58:15 2020  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_D\methods\Aromatic EPH 083120.M  
 Quant Title : GC Extractables  
 QLast Update : Tue Sep 01 02:30:57 2020  
 Response via : Initial Calibration  
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl  
 Signal Phase : Rxi-1ms  
 Signal Info : 20M x 0.18mm x 0.18µm

| Compound                           | R.T.  | Response | Conc Units    |
|------------------------------------|-------|----------|---------------|
| -----                              |       |          |               |
| System Monitoring Compounds        |       |          |               |
| 4) S 2-Bromonaphthalene (S...      | 7.685 | 10598666 | 103.334 ug/ml |
| Spiked Amount 50.000               |       | Recovery | = 206.67%     |
| 6) S 2-Fluorobiphenyl (SURR)       | 8.552 | 7338635  | 101.995 ug/ml |
| Spiked Amount 50.000 Range 0 - 131 |       | Recovery | = 203.99%#    |

Target Compounds

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_D\Data\FD092220AR\  
Data File : FD034805.D  
Signal(s) : FID2B.CH  
Acq On : 22 Sep 2020 12:46  
Operator : DD\AJ  
Sample : SURROGATE TEST  
Misc :  
ALS Vial : 54 Sample Multiplier: 1

Instrument :  
FID\_D  
ClientSampleId :  
SURROGATE TEST

Integration File: autoint1.e  
Quant Time: Sep 23 02:58:15 2020  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_D\methods\Aromatic EPH 083120.M  
Quant Title : GC Extractables  
QLast Update : Tue Sep 01 02:30:57 2020  
Response via : Initial Calibration  
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl  
Signal Phase : Rxi-1ms  
Signal Info : 20M x 0.18mm x 0.18µm

