

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD092721AL\
 Data File : FD038926.D
 Signal(s) : FID2B.CH
 Acq On : 28 Sep 2021 4:50
 Operator : AJ\MA
 Sample : M3943-04
 Misc :
 ALS Vial : 73 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Sep 28 06:47:20 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aliphatic EPH 083021.M
 Quant Title : GC Extractables
 QLast Update : Tue Aug 31 06:05:30 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.685	5652412	38.455 ug/ml
Spiked Amount 50.000		Recovery =	76.91%
12) S 1-chlorooctadecane (S...	13.112	4532520	38.966 ug/ml
Spiked Amount 50.000		Recovery =	77.93%
Target Compounds			
3) T A~Naphthalene (C11.7)	6.152f	246509	1.693 ug/ml
14) T n-Tetracosane (C24)	14.945	186273	1.394 ug/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD092721AL\
Data File : FD038926.D
Signal(s) : FID2B.CH
Acq On : 28 Sep 2021 4:50
Operator : AJ\MA
Sample : M3943-04
Misc :
ALS Vial : 73 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Sep 28 06:47:20 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aliphatic EPH 083021.M
Quant Title : GC Extractables
QLast Update : Tue Aug 31 06:05:30 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped


Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#3 A~Naphthalene (C11.7)

R.T.: 6.152 min
Delta R.T.: 0.054 min
Response: 246509
Conc: 1.69 ug/ml




#9 ortho-Terphenyl (SURR)

R.T.: 11.685 min
Delta R.T.: 0.008 min
Response: 5652412
Conc: 38.46 ug/ml



#12 1-chlorooctadecane (SURR)

R.T.: 13.112 min
Delta R.T.: 0.008 min
Response: 4532520
Conc: 38.97 ug/ml



#14 n-Tetracosane (C24)

R.T.: 14.945 min
Delta R.T.: -0.008 min
Response: 186273
Conc: 1.39 ug/ml