

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD101323AR\
 Data File : FD046753.D
 Signal(s) : FID2B.ch
 Acq On : 14 Oct 2023 02:11
 Operator : YP/AJ
 Sample : 04816-04
 Misc :
 ALS Vial : 72 Sample Multiplier: 1

Instrument :
 FID_D
 ClientSampleId :
 T-4

Integration File: autoint1.e
 Quant Time: Oct 14 03:03:30 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aromatic EPH 101023.M
 Quant Title : GC Extractables
 QLast Update : Wed Oct 11 02:49:06 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18µm

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
4) S 2-Bromonaphthalene (S...	7.491	12954717	69.786 ug/ml
Spiked Amount 50.000		Recovery =	139.57%
6) S 2-Fluorobiphenyl (SURR)	8.345	8597193	70.955 ug/ml
Spiked Amount 50.000 Range 0 - 131		Recovery =	141.91%#
11) S ortho-Terphenyl (SURR)	11.382	8810358	40.722 ug/ml
Spiked Amount 50.000		Recovery =	81.44%

Target Compounds

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD101323AR\
Data File : FD046753.D
Signal(s) : FID2B.ch
Acq On : 14 Oct 2023 02:11
Operator : YP/AJ
Sample : 04816-04
Misc :
ALS Vial : 72 Sample Multiplier: 1

Instrument :
FID_D
ClientSampleId :
T-4

Integration File: autoint1.e
Quant Time: Oct 14 03:03:30 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aromatic EPH 101023.M
Quant Title : GC Extractables
QLast Update : Wed Oct 11 02:49:06 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18µm






#4 2-Bromonaphthalene (SURRE)


R.T.: 7.491 min
Delta R.T.: 0.004 min
Response: 12954717
Conc: 69.79 ug/ml

Instrument :
FID_D
ClientSampleId :
T-4



#6 2-Fluorobiphenyl (SURRE)

R.T.: 8.345 min
Delta R.T.: 0.004 min
Response: 8597193
Conc: 70.96 ug/ml



#11 ortho-Terphenyl (SURRE)

R.T.: 11.382 min
Delta R.T.: 0.002 min
Response: 8810358
Conc: 40.72 ug/ml