

Comment	GC-EXT-FRZ#3
S/N	9246290
Sensor	Internal NTC (ML II)
Low	-40.00
Pre-low	
Pre-high	
High	-10.00

Index	Date and Time	Internal NTC (ML II) 9246290 GC-EXT-FRZ#3
1	18-10-21 12:00:00	-22.75°C
2	18-10-21 14:00:00	-22.43°C
3	18-10-21 16:00:00	-22.62°C
4	18-10-21 18:00:00	-24.71°C
5	18-10-21 20:00:00	-23.40°C
6	18-10-21 22:00:00	-22.32°C
7	19-10-21 00:00:00	H -4.30°C
8	19-10-21 02:00:00	-22.26°C
9	19-10-21 04:00:00	-23.16°C
10	19-10-21 06:00:00	-21.28°C
11	19-10-21 08:00:00	-21.52°C
12	19-10-21 10:00:00	-23.34°C
13	19-10-21 12:00:00	-23.72°C
14	19-10-21 14:00:00	-20.21°C