

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD103122AR\
 Data File : FD044368.D
 Signal(s) : FID2B.ch
 Acq On : 31 Oct 2022 18:20
 Operator : YP/AJ
 Sample : N5272-19RE
 Misc :
 ALS Vial : 66 Sample Multiplier: 1

Instrument :
 FID_D
 ClientSampleId :
 20221109-COMP-08RE

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/01/2022
 Supervised By :Ankita Jodhani 11/01/2022

Integration File: autoint1.e
 Quant Time: Oct 31 18:51:57 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aromatic EPH 101322.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 14 01:33:37 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18µm

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
4) S 2-Bromonaphthalene (S...	7.409	3564526	28.710 ug/mlm
Spiked Amount 50.000		Recovery =	57.42%
6) S 2-Fluorobiphenyl (SURR)	8.272	2021353	26.845 ug/ml
Spiked Amount 50.000 Range 0 - 131		Recovery =	53.69%
11) S ortho-Terphenyl (SURR)	11.291	3237611	19.128 ug/ml
Spiked Amount 50.000		Recovery =	38.26%

Target Compounds

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD103122AR\
Data File : FD044368.D
Signal(s) : FID2B.ch
Acq On : 31 Oct 2022 18:20
Operator : YP/AJ
Sample : N5272-19RE
Misc :
ALS Vial : 66 Sample Multiplier: 1


Instrument :
FID_D
ClientSampleId :
20221109-COMP-08RE


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/01/2022
Supervised By :Ankita Jodhani 11/01/2022

Integration File: autoint1.e
Quant Time: Oct 31 18:51:57 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aromatic EPH 101322.M
Quant Title : GC Extractables
QLast Update : Fri Oct 14 01:33:37 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18µm






#4 2-Bromonaphthalene (SURR)

R.T.: 7.409 min
Delta R.T.: -0.014 min
Response: 3564526
Conc: 28.71 ug/ml

Instrument :
FID_D
ClientSampleId :
20221109-COMP-08RE


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/01/2022
Supervised By :Ankita Jodhani 11/01/2022



#6 2-Fluorobiphenyl (SURR)

R.T.: 8.272 min
Delta R.T.: -0.014 min
Response: 2021353
Conc: 26.84 ug/ml



#11 ortho-Terphenyl (SURR)

R.T.: 11.291 min
Delta R.T.: -0.014 min
Response: 3237611
Conc: 19.13 ug/ml