

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD110424AR\
 Data File : FD048662.D
 Signal(s) : FID2B.ch
 Acq On : 05 Nov 2024 05:24
 Operator : YP/AJ
 Sample : 20 PPM AROMATIC HC STD
 Misc :
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 FID_D
 LabSampleId :
 20 PPM AROMATIC HC STD

Integration File: autoint1.e
 Quant Time: Nov 05 05:56:54 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aromatic EPH 102324.M
 Quant Title : GC Extractables
 QLast Update : Wed Oct 23 18:14:59 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound		AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 T	1,2,3-Trimethylbenzene (C10	173.971	180.978	E3	-4.0	104	0.00
2 T	Naphthalene (C11.7)	178.434	186.862	E3	-4.7	104	0.00
3 T	2-Methylnaphthalene (C12.89	178.865	186.350	E3	-4.2	104	0.00
4 S	2-Bromonaphthalene (SURR)	151.779	159.263	E3	-4.9	105	0.00
5 T	Acenaphthylene (C15.06)	175.592	179.936	E3	-2.5	102	0.00
6 S	2-Fluorobiphenyl (SURR)	98.192	103.040	E3	-4.9	103	0.00
7 T	Acenaphthene (C15.5)	177.363	184.785	E3	-4.2	104	0.00
8 T	Flouorene (C16.55)	169.037	178.078	E3	-5.3	105	0.00
9 T	Phenanthrene (C19.36)	165.306	173.398	E3	-4.9	104	0.00
10 T	Anthracene (C19.43)	159.910	131.569	E3	17.7	82	0.00
11 S	ortho-Terphenyl (SURR)	170.340	179.320	E3	-5.3	106	0.00
12 T	Fluoranthene (C21.85)	164.577	174.574	E3	-6.1	106	0.00
13 T	Pyrene (C20.8)	161.077	171.283	E3	-6.3	106	0.00
14 T	Benzo[a]anthracene (C26.37)	146.215	154.873	E3	-5.9	104	0.00
15 T	Chrysene (C27.41)	164.492	165.108	E3	-0.4	102	0.00
16 T	benzo[b]fluoranthene (C30.4	152.029	166.402	E3	-9.5	108	0.00
17 T	Bnezo[k]fluoranthene (C30.1	150.304	165.380	E3	-10.0	109	0.00
18 T	Benzo[a]pyrene (C31.34)	150.496	159.045	E3	-5.7	105	-0.01
19 T	Indeno[1,2,3-cd]pyrene (C35	130.478	143.405	E3	-9.9	109	-0.01
20 T	Dibenz[a,h]anthracene (C30.	162.735	189.590	E3	-16.5	116	-0.02
21 T	Benzo[g,h,i]perylene (C34.0	153.092	159.990	E3	-4.5	104	0.00

Evaluate Continuing Calibration Report - Not Found

(#) = Out of Range

SPCC's out = 0 CCC's out = 0