

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD110818AR\
 Data File : FD026438.D
 Signal(s) : FID2B.CH
 Acq On : 08 Nov 2018 19:36
 Operator : AJ\SJ
 Sample : 20 PPM AROMATIC HC STD
 Misc :
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 FID_D
 LabSampleId :
 20 PPM AROMATIC HC STD

Integration File: autoint1.e
 Quant Time: Nov 09 08:33:11 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aromatic EPH 110818.M
 Quant Title : GC Extractables
 QLast Update : Thu Nov 08 18:15:51 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 T	1,2,3-Trimethylbenzene (C10	79.823	79.088 E3	0.9	102	0.00
2 T	Naphthalene (C11.7)	87.761	87.616 E3	0.2	103	0.00
3 T	2-Methylnaphthalene (C12.89	86.673	87.105 E3	-0.5	103	0.00
4 S	2-Bromonaphthalene (SURR)	75.928	76.525 E3	-0.8	104	0.00
5 T	Acenaphthylene (C15.06)	84.761	85.480 E3	-0.8	103	0.00
6 S	2-Fluorobiphenyl (SURR)	52.289	52.562 E3	-0.5	102	0.00
7 T	Acenaphthene (C15.5)	88.413	89.148 E3	-0.8	103	0.00
8 T	Fluorene (C16.55)	88.131	88.213 E3	-0.1	101	0.00
9 T	Phenanthrene (C19.36)	87.789	86.253 E3	1.7	99	0.00
10 T	Anthracene (C19.43)	87.553	86.270 E3	1.5	99	0.00
11 S	ortho-Terphenyl (SURR)	84.761	83.123 E3	1.9	99	0.00
12 T	Fluoranthene (C21.85)	87.091	84.959 E3	2.4	98	0.00
13 T	Pyrene (C20.8)	86.675	85.261 E3	1.6	99	0.00
14 T	Benzo[a]anthracene (C26.37)	85.037	84.138 E3	1.1	100	0.00
15 T	Chrysene (C27.41)	85.697	84.648 E3	1.2	100	0.00
16 T	benzo[b]fluoranthene (C30.4	93.092	91.750 E3	1.4	100	0.00
17 T	Bnezo[k]fluoranthene (C30.1	92.602	91.813 E3	0.9	100	0.00
18 T	Benzo[a]pyrene (C31.34)	94.385	92.992 E3	1.5	100	0.00
19 T	Indeno[1,2,3-cd]pyrene (C35	96.027	96.638 E3	-0.6	102	0.00
20 T	Dibenz[a,h]anthracene (C30.	93.868	93.192 E3	0.7	100	0.00
21 T	Benzo[g,h,i]perylene (C34.0	98.311	96.391 E3	2.0	100	0.00

Evaluate Continuing Calibration Report - Not Found

(#) = Out of Range

SPCC's out = 0 CCC's out = 0