

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD110818AR\
 Data File : FD026451.D
 Signal(s) : FID2B.CH
 Acq On : 09 Nov 2018 7:56
 Operator : AJ\SJ
 Sample : 20 PPM AROMATIC HC STD
 Misc :
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 FID_D
 LabSampleId :
 20 PPM AROMATIC HC STD

Integration File: autoint1.e
 Quant Time: Nov 09 08:35:30 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aromatic EPH 110818.M
 Quant Title : GC Extractables
 QLast Update : Thu Nov 08 18:15:51 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 T	1,2,3-Trimethylbenzene (C10	79.823	79.871 E3	-0.1	103	0.00
2 T	Naphthalene (C11.7)	87.761	86.201 E3	1.8	101	0.00
3 T	2-Methylnaphthalene (C12.89	86.673	83.981 E3	3.1	100	0.00
4 S	2-Bromonaphthalene (SURR)	75.928	72.592 E3	4.4	98	0.00
5 T	Acenaphthylene (C15.06)	84.761	81.198 E3	4.2	98	0.00
6 S	2-Fluorobiphenyl (SURR)	52.289	50.057 E3	4.3	97	0.00
7 T	Acenaphthene (C15.5)	88.413	84.624 E3	4.3	98	0.00
8 T	Fluorene (C16.55)	88.131	83.657 E3	5.1	96	0.00
9 T	Phenanthrene (C19.36)	87.789	84.783 E3	3.4	97	0.00
10 T	Anthracene (C19.43)	87.553	84.524 E3	3.5	97	0.00
11 S	ortho-Terphenyl (SURR)	84.761	82.187 E3	3.0	98	0.00
12 T	Fluoranthene (C21.85)	87.091	89.459 E3	-2.7	104	0.00
13 T	Pyrene (C20.8)	86.675	91.000 E3	-5.0	106	0.00
14 T	Benzo[a]anthracene (C26.37)	85.037	93.890 E3	-10.4	112	0.00
15 T	Chrysene (C27.41)	85.697	94.771 E3	-10.6	112	0.00
16 T	benzo[b]fluoranthene (C30.4	93.092	98.622 E3	-5.9	108	0.00
17 T	Bnezo[k]fluoranthene (C30.1	92.602	98.374 E3	-6.2	108	0.00
18 T	Benzo[a]pyrene (C31.34)	94.385	98.393 E3	-4.2	106	0.00
19 T	Indeno[1,2,3-cd]pyrene (C35	96.027	98.734 E3	-2.8	104	0.00
20 T	Dibenz[a,h]anthracene (C30.	93.868	96.164 E3	-2.4	103	0.00
21 T	Benzo[g,h,i]perylene (C34.0	98.311	99.265 E3	-1.0	103	0.00

Evaluate Continuing Calibration Report - Not Found

(#) = Out of Range

SPCC's out = 0 CCC's out = 0