

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD111519AL\
 Data File : FD031531.D
 Signal(s) : FID2B.CH
 Acq On : 15 Nov 2019 19:25
 Operator : DD\AJ
 Sample : K5863-02
 Misc :
 ALS Vial : 63 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Nov 15 23:46:20 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Alaphatic EPH 110519.M
 Quant Title : GC Extractables
 QLast Update : Wed Nov 06 06:45:13 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.544	4649955	30.559 ug/ml
Spiked Amount 50.000		Recovery =	61.12%
12) S 1-chlorooctadecane (S...	12.973	3478689	30.613 ug/ml
Spiked Amount 50.000		Recovery =	61.23%
Target Compounds			
6) T n-Tetradecane (C14)	8.257f	204093	1.515 ug/ml
10) T n-Eicosane (C20)	12.581f	446954	3.388 ug/ml
11) T n-Heneicosane (C21)	13.199f	94176	0.712 ug/ml
14) T n-Tetracosane (C24)	14.785	299717	2.290 ug/ml
16) T n-Octacosane (C28)	16.775	129453	1.012 ug/ml
17) T n-Tricontane (C30)	17.660	163849	1.285 ug/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD111519AL\
Data File : FD031531.D
Signal(s) : FID2B.CH
Acq On : 15 Nov 2019 19:25
Operator : DD\AJ
Sample : K5863-02
Misc :
ALS Vial : 63 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Nov 15 23:46:20 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Alaphatic EPH 110519.M
Quant Title : GC Extractables
QLast Update : Wed Nov 06 06:45:13 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped


Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#6 n-Tetradecane (C14)

R.T.: 8.257 min
Delta R.T.: 0.064 min
Response: 204093
Conc: 1.51 ug/ml




#9 ortho-Terphenyl (SURR)

R.T.: 11.544 min
Delta R.T.: 0.009 min
Response: 464995
Conc: 30.56 ug/ml




#10 n-Eicosane (C20)

R.T.: 12.581 min
Delta R.T.: 0.056 min
Response: 446954
Conc: 3.39 ug/ml




#11 n-Heneicosane (C21)

R.T.: 13.199 min
Delta R.T.: 0.066 min
Response: 94176
Conc: 0.71 ug/ml




#12 1-chlorooctadecane (SURR)

R.T.: 12.973 min
Delta R.T.: 0.007 min
Response: 3478689
Conc: 30.61 ug/ml




#14 n-Tetracosane (C24)

R.T.: 14.785 min
Delta R.T.: -0.030 min
Response: 299717
Conc: 2.29 ug/ml



#16 n-Octacosane (C28)

R.T.: 16.775 min
Delta R.T.: 0.000 min
Response: 129453
Conc: 1.01 ug/ml



#17 n-Tricontane (C30)

R.T.: 17.660 min
Delta R.T.: 0.003 min
Response: 163849
Conc: 1.28 ug/ml