

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD111921AL\
 Data File : FD039534.D
 Signal(s) : FID2B.CH
 Acq On : 19 Nov 2021 16:46
 Operator : AJ\MA
 Sample : M4744-09DL 4X
 Misc :
 ALS Vial : 59 Sample Multiplier: 1

Instrument :
 FID_D
 ClientSampleId :
 B-4-15DL

Manual Integrations
 APPROVED

Reviewed By :Ankita Jodhani 11/22/2021
 Supervised By :Sohil Jodhani 11/22/2021

Integration File: autoint1.e
 Quant Time: Nov 19 21:35:39 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aliphatic EPH 102621.M
 Quant Title : GC Extractables
 QLast Update : Wed Oct 27 04:49:08 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.660	1118444	8.766 ug/ml
Spiked Amount 50.000		Recovery =	17.53%
12) S 1-chlorooctadecane (S...	13.087	867190	8.344 ug/mlm
Spiked Amount 50.000		Recovery =	16.69%

Target Compounds

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD111921AL\
Data File : FD039534.D
Signal(s) : FID2B.CH
Acq On : 19 Nov 2021 16:46
Operator : AJ\MA
Sample : M4744-09DL 4X
Misc :
ALS Vial : 59 Sample Multiplier: 1


Instrument :
FID_D
ClientSampleId :
B-4-15DL


Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 11/22/2021
Supervised By :Sohil Jodhani 11/22/2021

Integration File: autoint1.e
Quant Time: Nov 19 21:35:39 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aliphatic EPH 102621.M
Quant Title : GC Extractables
QLast Update : Wed Oct 27 04:49:08 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.660 min
Delta R.T.: -0.009 min
Response: 1118444
Conc: 8.77 ug/ml

Instrument :
FID_D
ClientSampleId :
B-4-15DL

Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 11/22/2021
Supervised By :Sohil Jodhani 11/22/2021



#12 1-chlorooctadecane (SURR)

R.T.: 13.087 min
Delta R.T.: -0.009 min
Response: 867190
Conc: 8.34 ug/ml m