

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD111921AL\
 Data File : FD039563.D
 Signal(s) : FID2B.CH
 Acq On : 20 Nov 2021 11:26
 Operator : AJ\MA
 Sample : M4758-11
 Misc :
 ALS Vial : 86 Sample Multiplier: 1

Instrument :
 FID_D
 ClientSampleId :
 TP6

Manual Integrations
 APPROVED

Reviewed By :Ankita Jodhani 11/22/2021
 Supervised By :Sohil Jodhani 11/22/2021

Integration File: autoint1.e
 Quant Time: Nov 21 23:59:39 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aliphatic EPH 102621.M
 Quant Title : GC Extractables
 QLast Update : Wed Oct 27 04:49:08 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.657	3619566	28.369 ug/ml
Spiked Amount 50.000		Recovery =	56.74%
12) S 1-chlorooctadecane (S...	13.085	2944765	28.335 ug/mlm
Spiked Amount 50.000		Recovery =	56.67%

Target Compounds

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD111921AL\
Data File : FD039563.D
Signal(s) : FID2B.CH
Acq On : 20 Nov 2021 11:26
Operator : AJ\MA
Sample : M4758-11
Misc :
ALS Vial : 86 Sample Multiplier: 1


Instrument :
FID_D
ClientSampleId :
TP6


Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 11/22/2021
Supervised By :Sohil Jodhani 11/22/2021

Integration File: autoint1.e
Quant Time: Nov 21 23:59:39 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aliphatic EPH 102621.M
Quant Title : GC Extractables
QLast Update : Wed Oct 27 04:49:08 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.657 min
Delta R.T.: -0.012 min
Response: 3619566
Conc: 28.37 ug/ml

Instrument :
FID_D
ClientSampleId :
TP6

Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 11/22/2021
Supervised By :Sohil Jodhani 11/22/2021



#12 1-chlorooctadecane (SURR)

R.T.: 13.085 min
Delta R.T.: -0.012 min
Response: 2944765
Conc: 28.34 ug/ml m