

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD121619AL\
 Data File : FD031820.D
 Signal(s) : FID2B.CH
 Acq On : 16 Dec 2019 20:40
 Operator : DD\AJ
 Sample : K6113-08
 Misc :
 ALS Vial : 59 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Dec 17 01:27:02 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Alaphatic EPH 120619.M
 Quant Title : GC Extractables
 QLast Update : Fri Dec 06 04:20:46 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.542	4884828	33.771 ug/ml
Spiked Amount 50.000		Recovery =	67.54%
12) S 1-chlorooctadecane (S...	12.971	3959553	34.916 ug/ml
Spiked Amount 50.000		Recovery =	69.83%
Target Compounds			
6) T n-Tetradecane (C14)	8.256	217066	1.654 ug/ml
10) T n-Eicosane (C20)	12.580	765752	5.722 ug/ml
14) T n-Tetracosane (C24)	14.783	302248	2.253 ug/ml
16) T n-Octacosane (C28)	16.770	107904	0.851 ug/ml
17) T n-Tricontane (C30)	17.655	117610	0.993 ug/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD121619AL\
Data File : FD031820.D
Signal(s) : FID2B.CH
Acq On : 16 Dec 2019 20:40
Operator : DD\AJ
Sample : K6113-08
Misc :
ALS Vial : 59 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Dec 17 01:27:02 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Alaphatic EPH 120619.M
Quant Title : GC Extractables
QLast Update : Fri Dec 06 04:20:46 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped


Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#6 n-Tetradecane (C14)

R.T.: 8.256 min
Delta R.T.: 0.047 min
Response: 217066
Conc: 1.65 ug/ml




#9 ortho-Terphenyl (SURR)

R.T.: 11.542 min
Delta R.T.: -0.013 min
Response: 4884828
Conc: 33.77 ug/ml




#10 n-Eicosane (C20)

R.T.: 12.580 min
Delta R.T.: 0.040 min
Response: 765752
Conc: 5.72 ug/ml




#12 1-chlorooctadecane (SURR)

R.T.: 12.971 min
Delta R.T.: -0.011 min
Response: 3959553
Conc: 34.92 ug/ml




#14 n-Tetracosane (C24)

R.T.: 14.783 min
Delta R.T.: -0.046 min
Response: 302248
Conc: 2.25 ug/ml



#16 n-Octacosane (C28)

R.T.: 16.770 min
Delta R.T.: -0.017 min
Response: 107904
Conc: 0.85 ug/ml



#17 n-Tricontane (C30)

R.T.: 17.655 min
Delta R.T.: -0.014 min
Response: 117610
Conc: 0.99 ug/ml