

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_D\Data\FD121721AL\  
 Data File : FD040006.D  
 Signal(s) : FID2B.CH  
 Acq On : 17 Dec 2021 19:34  
 Operator : AJ\MA  
 Sample : M5105-01  
 Misc :  
 ALS Vial : 67 Sample Multiplier: 1

Instrument :  
 FID\_D  
 ClientSampleId :  
 HD-01-121521

Manual Integrations  
 APPROVED

Reviewed By :Yogesh Patel 12/20/2021  
 Supervised By :Ankita Jodhani 12/23/2021

Integration File: autoint1.e  
 Quant Time: Dec 17 23:45:00 2021  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_D\methods\Aliphatic EPH 121521.M  
 Quant Title : GC Extractables  
 QLast Update : Wed Dec 15 16:44:55 2021  
 Response via : Initial Calibration  
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul  
 Signal Phase : Rxi-1ms  
 Signal Info : 20M x 0.18mm x 0.18um

| Compound                       | R.T.   | Response | Conc Units    |
|--------------------------------|--------|----------|---------------|
| -----                          |        |          |               |
| System Monitoring Compounds    |        |          |               |
| 9) S ortho-Terphenyl (SURR)    | 10.760 | 5856789  | 37.307 ug/mlm |
| Spiked Amount                  | 50.000 | Recovery | = 74.61%      |
| 12) S 1-chlorooctadecane (S... | 12.216 | 64920090 | 525.697 ug/ml |
| Spiked Amount                  | 50.000 | Recovery | = 1051.39%    |

Target Compounds

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_D\Data\FD121721AL\  
Data File : FD040006.D  
Signal(s) : FID2B.CH  
Acq On : 17 Dec 2021 19:34  
Operator : AJ\MA  
Sample : M5105-01  
Misc :  
ALS Vial : 67 Sample Multiplier: 1

Instrument :  
FID\_D  
ClientSampleId :  
HD-01-121521

Manual Integrations  
APPROVED

Reviewed By :Yogesh Patel 12/20/2021  
Supervised By :Ankita Jodhani 12/23/2021

Integration File: autoint1.e  
Quant Time: Dec 17 23:45:00 2021  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_D\methods\Aliphatic EPH 121521.M  
Quant Title : GC Extractables  
QLast Update : Wed Dec 15 16:44:55 2021  
Response via : Initial Calibration  
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul  
Signal Phase : Rxi-1ms  
Signal Info : 20M x 0.18mm x 0.18um

