

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD123021AL\
 Data File : FD040330.D
 Signal(s) : FID2B.CH
 Acq On : 31 Dec 2021 8:05
 Operator : AJ\MA
 Sample : M5255-01
 Misc :
 ALS Vial : 72 Sample Multiplier: 1

Instrument :
 FID_D
 ClientSampleId :
 TR-04-122921

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 01/03/2022
 Supervised By :Ankita Jodhani 01/03/2022

Integration File: autoint1.e
 Quant Time: Dec 31 08:21:57 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aliphatic EPH 122821.M
 Quant Title : GC Extractables
 QLast Update : Wed Dec 29 02:26:30 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	10.691	4484527	28.786 ug/mlm
Spiked Amount	50.000	Recovery	= 57.57%
12) S 1-chlorooctadecane (S...	12.170	134155750	1112.415 ug/ml
Spiked Amount	50.000	Recovery	= 2224.83%

Target Compounds

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD123021AL\
Data File : FD040330.D
Signal(s) : FID2B.CH
Acq On : 31 Dec 2021 8:05
Operator : AJ\MA
Sample : M5255-01
Misc :
ALS Vial : 72 Sample Multiplier: 1


Instrument :
FID_D
ClientSampleId :
TR-04-122921


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 01/03/2022
Supervised By :Ankita Jodhani 01/03/2022

Integration File: autoint1.e
Quant Time: Dec 31 08:21:57 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aliphatic EPH 122821.M
Quant Title : GC Extractables
QLast Update : Wed Dec 29 02:26:30 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 10.691 min
Delta R.T.: -0.041 min
Response: 4484527
Conc: 28.79 ug/ml

Instrument :
FID_D
ClientSampleId :
TR-04-122921

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 01/03/2022
Supervised By :Ankita Jodhani 01/03/2022



#12 1-chlorooctadecane (SURR)

R.T.: 12.170 min
Delta R.T.: 0.007 min
Response: 134155750
Conc: 1112.42 ug/ml