

Method Path : P:\HPCHEM1\FID_D\methods\
 Method File : FD010318.M
 Title :
 Last Update : Wed Jan 03 14:14:22 2018
 Response Via : Initial Calibration

Calibration Files

50 =FD018964.D 100 =FD018963.D 20 =FD018965.D
 10 =FD018966.D 5 =FD018967.D

	Compound	50	100	20	10	5	Avg		%RSD
1)	N-OCTANE	8.010	8.005	6.888	8.073	7.870	7.769	E4	6.41
2)	N-DECANE	8.366	8.209	6.909	8.190	8.047	7.944	E4	7.42
3)	N-DODECANE	8.753	8.498	7.092	8.400	8.223	8.193	E4	7.87
4)	N-TETRADECANE	9.195	8.753	7.419	8.732	8.592	8.538	E4	7.79
5)	N-HEXADECANE	9.307	8.740	7.626	9.010	8.824	8.702	E4	7.35
6)	N-OCTADECANE	9.378	8.753	7.880	9.439	9.510	8.992	E4	7.69
7)	N-EICOSANE	9.149	8.582	7.736	9.282	9.288	8.807	E4	7.56
8)	N-DOCOSANE	8.992	8.525	7.536	9.081	8.997	8.626	E4	7.50
9) S	TETRACOSANE-d50 (S...	7.527	7.208	6.370	7.708	7.531	7.269	E4	7.34
10)	N-TETRACOSANE	8.928	8.606	7.438	9.043	8.986	8.600	E4	7.81
11)	N-HEXACOSANE	8.885	8.740	7.366	9.003	8.980	8.595	E4	8.08
12)	N-OCTACOSANE	8.979	8.924	7.413	9.154	9.053	8.705	E4	8.35
13)	N-TRIACONTANE	9.341	9.307	7.752	9.602	9.681	9.137	E4	8.65
14)	N-DOTRIACONTANE	0.958	0.952	0.829	1.010	1.019	0.953	E5	7.96
15)	N-TETRATRIACONTANE	9.233	9.142	7.896	9.734	9.652	9.131	E4	8.07
16)	N-HEXATRIACONTANE	8.120	8.077	6.700	8.375	8.211	7.897	E4	8.59
17)	N-OCTATRIACONTANE	8.100	8.021	6.917	8.465	8.487	7.998	E4	8.00
18)	N-TETRACONTANE	8.139	8.070	7.287	8.534	8.368	8.080	E4	5.94

(#) = Out of Range