

Data Path : P:\HPCHEM1\FID D\Data\FD013018\
 Data File : FD019303.D
 Signal(s) : FID2B.CH
 Acq On : 30 Jan 2018 9:52
 Operator : UA/AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jan 30 11:53:22 2018
 Quant Method : P:\HPCHEM1\FID_D\methods\FD010318.M
 Quant Title :
 QLast Update : Wed Jan 03 14:14:22 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

	Compound	R.T.	Response	Conc Units

System Monitoring Compounds				
9) S	TETRACOSANE-d50 (SURR...	14.341	3876293	53.329 ug/ml
Target Compounds				
2)	N-DECANE	4.135	3787979	47.681 ug/ml
3)	N-DODECANE	6.142	3872607	47.265 ug/ml
4)	N-TETRADECANE	7.926	3995275	46.792 ug/ml
5)	N-HEXADECANE	9.517	4108539	47.216 ug/ml
6)	N-OCTADECANE	10.950	4283622	47.638 ug/ml
7)	N-EICOSANE	12.254	4382909	49.764 ug/ml
8)	N-DOCOSANE	13.445	4520160	52.400 ug/ml
10)	N-TETRACOSANE	14.543	4689180	54.525 ug/ml
11)	N-HEXACOSANE	15.559	4752933	55.300 ug/ml
12)	N-OCTACOSANE	16.504	4768415	54.780 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : P:\HPCHEM1\FID D\Data\FD013018\
Data File : FD019303.D
Signal(s) : FID2B.CH
Acq On : 30 Jan 2018 9:52
Operator : UA/AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jan 30 11:53:22 2018
Quant Method : P:\HPCHEM1\FID_D\methods\FD010318.M
Quant Title :
QLast Update : Wed Jan 03 14:14:22 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

