

Data Path : P:\HPCHEM1\FID_D\Data\FD052118AR\
 Data File : FD023782.D
 Signal(s) : FID2B.CH
 Acq On : 22 May 2018 4:22
 Operator : UA/AJ
 Sample : 20 PPM AROMATIC HC STD
 Misc :
 ALS Vial : 54 Sample Multiplier: 1

Integration File: sample.E
 Quant Time: May 22 17:08:56 2018
 Quant Method : P:\HPCHEM1\FID_D\methods\Aromatic EPH 052118.M
 Quant Title : GC Extractables
 QLast Update : Tue May 22 10:27:34 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 T	1,2,3-Trimethylbenzene (C10	86.744	87.532 E3	-0.9	105	0.00
2 T	Naphthalene (C11.7)	91.972	92.079 E3	-0.1	105	0.00
3 T	2-Methylnaphthalene (C12.89	90.566	90.294 E3	0.3	105	0.00
4 S	2-Bromonaphthalene (SURR)	80.302	80.212 E3	0.1	105	0.00
5 T	Acenaphthylene (C15.06)	88.629	88.443 E3	0.2	105	0.00
6 S	2-Fluorobiphenyl (SURR)	57.058	57.004 E3	0.1	105	0.00
7 T	Acenaphthene (C15.5)	92.762	92.617 E3	0.2	105	0.00
8 T	Flouorene (C16.55)	93.450	93.292 E3	0.2	105	0.00
9 T	Phenanthrene (C19.36)	97.925	97.628 E3	0.3	105	0.00
10 T	Anthracene (C19.43)	97.462	97.103 E3	0.4	105	0.00
11 S	ortho-Terphenyl (SURR)	92.984	92.677 E3	0.3	105	0.00
12 T	Fluoranthene (C21.85)	102.231	101.747 E3	0.5	103	0.00
13 T	Pyrene (C20.8)	103.433	102.644 E3	0.8	104	0.00
14 T	Benzo[a]anthracene (C26.37)	100.811	99.677 E3	1.1	103	0.00
15 T	Chrysene (C27.41)	101.927	101.173 E3	0.7	104	0.00
16 T	benzo[b]fluoranthene (C30.4	100.976	99.407 E3	1.6	102	0.00
17 T	Bnezo[k]fluoranthene (C30.1	102.437	100.794 E3	1.6	103	0.00
18 T	Benzo[a]pyrene (C31.34)	98.481	96.355 E3	2.2	102	0.00
19 T	Indeno[1,2,3-cd]pyrene (C35	82.123	72.882 E3	11.3	90	0.00
20 T	Dibenz[a,h]anthracene (C30.	90.055	84.246 E3	6.5	95	0.00
21 T	Benzo[g,h,i]perylene (C34.0	83.352	71.504 E3	14.2	83	0.00

Evaluate Continuing Calibration Report - Not Founds

(#) = Out of Range

SPCC's out = 0 CCC's out = 0