

Method Path : P:\HPCHEM1\FID_E\methods\
 Method File : FE011218.M
 Title :
 Last Update : Fri Jan 12 18:22:53 2018
 Response Via : Initial Calibration

Calibration Files

50 =FE025514.D 100 =FE025513.D 20 =FE025515.D
 10 =FE025516.D 5 =FE025517.D

	Compound	50	100	20	10	5	Avg		%RSD
1)	N-OCTANE	1.468	1.378	1.424	1.492	1.271	1.407	E5	6.21
2)	N-DECANE	1.594	1.504	1.519	1.551	1.295	1.492	E5	7.74
3)	N-DODECANE	1.758	1.701	1.627	1.580	1.311	1.595	E5	10.86
4)	N-TETRADECANE	1.957	1.909	1.870	1.777	1.600	1.822	E5	7.74
5)	N-HEXADECANE	2.085	2.020	2.114	2.113	1.882	2.043	E5	4.78
6)	N-OCTADECANE	2.203	2.104	2.258	2.281	2.205	2.210	E5	3.10
7)	N-EICOSANE	2.202	2.140	2.287	2.259	2.200	2.218	E5	2.58
8)	N-DOCOSANE	2.187	2.195	2.292	2.155	2.141	2.194	E5	2.70
9) S	TETRACOSANE-d50 (S...	1.810	1.856	1.939	1.664	1.825	1.819	E5	5.50
10)	N-TETRACOSANE	2.137	2.186	2.289	1.958	2.093	2.133	E5	5.72
11)	N-HEXACOSANE	1.958	2.006	2.143	1.658	1.871	1.927	E5	9.34
12)	N-OCTACOSANE	1.692	1.696	1.838	1.507	1.674	1.681	E5	6.99
13)	N-TRIACONTANE	1.613	1.582	1.642	1.538	1.519	1.579	E5	3.23
14)	N-DOTRIACONTANE	1.654	1.617	1.677	1.574	1.588	1.622	E5	2.68
15)	N-TETRATRIACONTANE	1.562	1.523	1.603	1.524	1.556	1.554	E5	2.13
16)	N-HEXATRIACONTANE	1.349	1.302	1.381	1.284	1.329	1.329	E5	2.89
17)	N-OCTATRIACONTANE	1.306	1.275	1.321	1.236	1.217	1.271	E5	3.49
18)	N-TETRACONTANE	1.327	1.284	1.345	1.224	1.226	1.281	E5	4.35

(#) = Out of Range