

rters

Instrument :
FID_E
LabSampleId :
20 PPM ALIPHATIC HC STD
Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 01/05/2023
Supervised By :Ankita Jodhani 01/05/2023

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE01042
Data File : FE039625.D
Signal(s) : FID1B.ch
Acq On : 04 Jan 2023 09:42
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 121722.M
Title : GC Extractables

Signal : FID1B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.319	3.281	3.372	BB	244633	2431557	78.02%	4.061%
2	4.569	4.530	4.621	BB	247381	2443159	78.39%	4.080%
3	6.273	6.236	6.367	BB	277351	2791815	89.58%	4.663%
4	6.734	6.694	6.784	BB	252634	2480600	79.59%	4.143%
5	7.368	7.320	7.468	BB	250989	2640146	84.71%	4.409%
6	8.559	8.518	8.621	BB	245093	2552377	81.90%	4.263%
7	10.166	10.118	10.231	BB	247846	2695704	86.50%	4.502%
8	11.608	11.514	11.664	BB	244371	2811467	90.21%	4.696%
9	11.926	11.881	11.986	BB	281053	3061897	98.24%	5.114%
10	12.917	12.868	12.972	BB	246942	2774236	89.01%	4.633%
11	13.361	13.314	13.408	BB	200834	2353016	75.50%	3.930%
12	13.528	13.497	13.577	M	236650	2791311	89.56%	4.662%
13	14.115	14.085	14.169	VB	232321	2803099	89.94%	4.682%
14	15.218	15.187	15.271	VB	226481	2833065	90.90%	4.732%
15	16.239	16.200	16.290	BB	210037	2799397	89.82%	4.675%
16	17.189	17.130	17.246	BB	220401	2883408	92.52%	4.816%
17	18.077	18.040	18.120	M	228447	3061279	98.23%	5.113%
18	18.911	18.841	18.964	BB	219921	3116598	100.00%	5.205%
19	19.694	19.651	19.736	BB	199840	2890516	92.75%	4.828%
20	20.435	20.401	20.471	BV	176118	2608603	83.70%	4.357%
21	21.170	21.101	21.251	BB	140955	2563039	82.24%	4.281%
22	22.081	22.020	22.157	BV	106706	2487971	79.83%	4.155%
Sum of corrected areas:						59874261		

Aliphatic EPH 121722.M Thu Jan 05 02:42:40 2023